



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96906

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	07/26/2018	16:40
2	CAL2	1	Water			21.7	7.00	07/26/2018	16:42
3	CAL3	1	Water			22.0	10.04	07/26/2018	16:44
4	ICV	1	Water			21.8	7.01	07/26/2018	16:46
5	CCV1	1	Water			21.7	2.00	07/26/2018	17:10
6	J4176-02	1	Water			23.9	6.54	07/26/2018	17:12
7	J4176-02DUP	1	Water			24.0	6.55	07/26/2018	17:14
8	J4188-02	1	Water			24.2	8.07	07/26/2018	17:16
9	J4189-27	1	Water			16.3	6.47	07/26/2018	17:18
10	CCV2	1	Water			21.9	12.00	07/26/2018	17:20

WORKLIST(Hardcopy Internal Chain)

LB 96906

WorkList Name : corrosivity072618-1

WorkList ID : 114935

Date : 7/26/2018 4:10:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/26/2018	Water	J4176-02	Corrosivity	Cool 4 deg C	PSEG01	K63	SV26835	07/26/2018	9040C
07/26/2018	Water	J4188-02	Corrosivity	Cool 4 deg C	PSEG01	K63	HA-BASELINEWATER-N4897	07/26/2018	9040C
07/26/2018	Water	J4189-27	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-12	07/26/2018	9040C

Date/Time 07/26/18 4:30

Received by: SP

Relinquished by: CP

Date/Time 07/26/18 5:30

Received by: CP

Relinquished by: SP