


ISM02.4

LB 96910

INSTRUMENT ID: CV4

Linear



A= 0.0000e+000

B= 2.4753e-004

C= -3.0825e-002

Rho= 0.9997435

Accept=Accepted

slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.004	-0.004	109	0.000	109					-
0.2	0.200	0.222	0.022	1020	0.0 %	1020					11
2.5	2.500	2.526	0.026	10331	0.0 %	10331					1
5.0	5.000	5.019	0.019	20402	0.0 %	20402					0
7.5	7.500	7.330	-0.170	29737	0.0 %	29737					-2
10.0	10.000	10.107	0.107	40956	0.0 %	40956					1

13
7/26/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	109	-		0 ISM02.4	PPB	7/26/2018 18:51	Std
	0.2 S0.2	1020	-		0.2 ISM02.4	PPB	7/26/2018 18:53	Std
	2.5 S2.5	10331	-		2.5 ISM02.4	PPB	7/26/2018 18:55	Std
	5 S5	20402	-		5 ISM02.4	PPB	7/26/2018 18:58	Std
	7.5 S7.5	29737	-		7.5 ISM02.4	PPB	7/26/2018 19:00	Std
	10 S10.0	40956	-		10 ISM02.4	PPB	7/26/2018 19:06	Std
ICV	ICV	15401	3.7813 -		ISM02.4	PPB	7/26/2018 19:09	SMPL
ICB	ICB	71	-0.0133 -		ISM02.4	PPB	7/26/2018 19:12	SMPL
CCV023	CCV023	20162	4.9598 -		ISM02.4	PPB	7/26/2018 19:14	SMPL
CCB023	CCB023	-17	-0.035 -		ISM02.4	PPB	7/26/2018 19:16	SMPL
PB111505BL	PBW003	242	0.0291 -		ISM02.4	PPB	7/26/2018 19:18	SMPL
J4174-01	MC0AD0	98	-0.0066 -		ISM02.4	PPB	7/26/2018 19:21	SMPL
J4174-02	MC0AD1	75	-0.0123 -		ISM02.4	PPB	7/26/2018 19:23	SMPL
J4174-03	MC0AD2	15	-0.0271 -		ISM02.4	PPB	7/26/2018 19:25	SMPL
J4174-04	MC0AD3	33	-0.0227 -		ISM02.4	PPB	7/26/2018 19:27	SMPL
J4174-05	MC0AD4	80	-0.011 -		ISM02.4	PPB	7/26/2018 19:30	SMPL
J4174-06	MC0AD4D	5	-0.0296 -		ISM02.4	PPB	7/26/2018 19:32	SMPL
J4174-07	MC0AD4S	4141	0.9942 -		ISM02.4	PPB	7/26/2018 19:34	SMPL
J4174-08	MC0AD5	120	-0.0011 -		ISM02.4	PPB	7/26/2018 19:37	SMPL
CCV024	CCV024	20395	5.0175 -		ISM02.4	PPB	7/26/2018 19:39	SMPL
CCB024	CCB024	22	-0.0254 -		ISM02.4	PPB	7/26/2018 19:41	SMPL