



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB96918
Analysis Date: 07/27/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 07/27/2018
Extraction IN Time: 07:45
Extraction OUT Time: 08:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB96918BL	LB96918BL	WATER	2	1000	100	3.0514	3.0514	3.02	3.0514	3.0514	0.0000	0
2	LB96918BS	LB96918BS	WATER	2	1000	100	3.0014	3.0014	3.01	3.0198	3.0198	0.0184	18.4
3	LB96918BSD	LB96918BSD	WATER	2	1000	100	3.1699	3.1699	3.04	3.1886	3.1886	0.0187	18.7
4	J4176-01	SV26835	WATER	2	1000	100	3.0313	3.0313	3.05	3.0332	3.0332	0.0019	1.9
5	J4188-01	HA-BASELINEWATER-N4897	WATER	2	1000	100	3.0094	3.0094	3.06	3.0122	3.0122	0.0028	2.8
6	J4189-26	TP-12	WATER	2	700	100	3.0939	3.0939	3.04	3.0951	3.0951	0.0012	1.71



QC Batch# LB96918 Test: TPH Analysis Date: 07/27/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2392
pH Paper 0-14	W2373
Sodium Sulfate	EP1866
1:1 HCL	WP64228
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	10.00 ML	WP64230
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>10:16</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>09:15</u>		
Bal Check Time:	<u>08:10</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:40</u>
		Out Time:	<u>10:15</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:51</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>11:15</u>		
Bal Check Time:	<u>13:00</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>12:25</u>
		Out Time:	<u>11:50</u>		

2896918

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH J4189

WorkList ID : 114946

Date : 7/27/2018 7:35:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/31/2018	Water	J4176-01	-1-TPH	Conc H2SO4 tc	PSEG01	K63	SV26835	07/26/2018	1664A
07/31/2018	Water	J4188-01	-1-TPH	Conc H2SO4 tc	PSEG01	K63	HA-BASELINEWATER-N4897	07/26/2018	1664A
07/31/2018	Water	J4189-26	-1-TPH	Conc H2SO4 tc	PSEG01	K63	TP-12	07/26/2018	1664A

Date/Time 07/27/18 8:40 AM
Received by: So
Relinquished by: So

Date/Time 07/27/18 10:21 AM
Received by: So
Relinquished by: So