

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB96921

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP66402
calibration std. phosphate 0.5 ppm	WP66401
calibration std. phosphate 0.3 ppm	WP66400
calibration std. 0 ppm	WP66397
phosphate ICV std.	WP66404
phosphate CCV std.	WP66403
calibration std. phosphate 0.1 ppm	WP66399
calibration std. phosphate 0.05 ppm	WP66398
Combined reagent	WP66484
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	07/27/2018	10:31
2	CAL2	0.05	1	50	50	0.033	07/27/2018	10:31
3	CAL3	0.10	1	50	50	0.065	07/27/2018	10:32
4	CAL4	0.30	1	50	50	0.186	07/27/2018	10:32
5	CAL5	0.50	1	50	50	0.330	07/27/2018	10:33
6	CAL6	1.00	1	50	50	0.649	07/27/2018	10:33



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB96921

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.500	07/27/2018	10:34
2	ICB		1	50	50	0.000	0.001	07/27/2018	10:34
3	CCV1	0.50	1	50	50	0.327	0.504	07/27/2018	10:35
4	CCB1		1	50	50	0.000	0.001	07/27/2018	10:35
5	PB111532BL		1	50	50	0.000	0.001	07/27/2018	10:36
6	PB111532BS	0.50	1	50	50	0.321	0.495	07/27/2018	10:37
7	J4183-01		1	50	50	0.324	0.500	07/27/2018	10:37
8	J4183-01DUP		1	50	50	0.345	0.532	07/27/2018	10:38
9	J4183-01MS	0.50	1	50	50	0.670	1.032	07/27/2018	10:38
10	J4183-01MSD	0.50	1	50	50	0.675	1.039	07/27/2018	10:39
11	J4183-01MS	0.50	2	50	50	0.305	0.470	07/27/2018	10:45
12	J4183-01MSD	0.50	2	50	50	0.301	0.464	07/27/2018	10:45
13	CCV2	0.50	1	50	50	0.326	0.503	07/27/2018	10:46
14	CCB2		1	50	50	0.000	0.001	07/27/2018	10:46