



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96925

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.9	07/27/2018	10:45
2	J4176-02		1	7	>100.0	>212.0	>212.0	07/27/2018	11:20
3	J4176-02DUP		1	7	>100.0	>212.0	>212.0	07/27/2018	11:50
4	J4188-02		1	8	>100.0	>212.0	>212.0	07/27/2018	12:15
5	J4189-27		1	8	>100.0	>212.0	>212.0	07/27/2018	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP72718

WorkList ID : 114979

Date : 7/27/2018 10:08:24 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/31/2018	Water	J4176-02	Flash Point	Cool 4 deg C	PSEG01	K63	SV26835	07/26/2018	1010A
07/31/2018	Water	J4188-02	Flash Point	Cool 4 deg C	PSEG01	K63	HA-BASELINEWATER-N4897	07/26/2018	1010A
07/31/2018	Water	J4189-27	Flash Point	Cool 4 deg C	PSEG01	K63	TP-12	07/26/2018	1010A

Date/Time 07/27/18 10:15 AM
 Received by: KS
 Relinquished by: CP

Date/Time 07/27/18 12:45 PM
 Received by: CP
 Relinquished by: KS

LB96925