



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96930

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/27/2018	09:00
2	CAL2	1	Water			21.8	7.00	07/27/2018	09:02
3	CAL3	1	Water			21.6	10.04	07/27/2018	09:04
4	ICV	1	Water			21.8	7.00	07/27/2018	09:06
5	CCV1	1	Water			21.6	2.00	07/27/2018	09:08
6	J4193-01	1	Solid	20.02	20	24.4	5.72	07/27/2018	09:10
7	J4193-01DUP	1	Solid	20.03	20	24.5	5.73	07/27/2018	09:12
8	CCV2	1	Water			21.8	12.00	07/27/2018	09:14

LB96930

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH072718

WorkList ID : 114960

Date : 7/27/2018 8:18:41 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/26/2018	Solid	J4193-01	pH	Cool 4 deg C	TULL02	N41	FK-PS-01-035	07/26/2018	9045C

Date/Time 07/27/18 8:30
Received by: SP
Relinquished by: CP

Date/Time 07/27/18 9:25
Received by: CP
Relinquished by: SP