

LB96938


245.1

INSTRUMENT ID: CX1

Linear

 μ Abs.:

33504



A= 0.0000e+000

B= 2.9707e-004

C= 7.2349e-003

Rho= 0.9999240

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	-62	0.000	-62				
0.05	0.050	0.046	-0.004	132	0.0 %	132				
0.2	0.200	0.225	0.025	733	0.0 %	733				
2.5	2.500	2.500	-0.000	8390	0.0 %	8390				
5.0	5.000	4.938	-0.062	16598	0.0 %	16598				
7.5	7.500	7.592	0.092	25532	0.0 %	25532				
10.0	10.000	9.960	-0.040	33504	0.0 %	33504				

LB96938 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Jul 2018 14:20:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-62	-	0.0000	245.1	PPB	27 Jul 2018 14:24:02	S	Std
0.05 - 1	50.05	132	-	0.0500	245.1	PPB	27 Jul 2018 14:26:10	S	Std
0.2 - 1	50.2	733	-	0.2000	245.1	PPB	27 Jul 2018 14:28:18	S	Std
2.5 - 1	52.5	8390	-	2.5000	245.1	PPB	27 Jul 2018 14:30:27	S	Std
5.0 - 1	55	16598	-	5.0000	245.1	PPB	27 Jul 2018 14:32:35	S	Std
7.5 - 1	57.5	25532	-	7.5000	245.1	PPB	27 Jul 2018 14:34:42	S	Std
10.0 - 1	316	33504	-	10.0000	245.1	PPB	27 Jul 2018 14:39:21	S	Std
ICV75 - 1	ICV75	13294	3.9565	-245.1		PPB	27 Jul 2018 14:41:56	U	SMPL
ICB75 - 1	ICB75	-223	-0.0590	-245.1		PPB	27 Jul 2018 14:44:03	U	SMPL
CCV31 - 1	CCV31	17036	5.0681	-245.1		PPB	27 Jul 2018 14:46:10	U	SMPL
CCB31 - 1	CCB31	-205	-0.0537	-245.1		PPB	27 Jul 2018 14:48:16	U	SMPL
CRA - 1	CRA	119	0.0426	-245.1		PPB	27 Jul 2018 14:50:24	U	SMPL
HighStd - 1	HighStd	33351	9.9147	-245.1		PPB	27 Jul 2018 14:52:30	U	SMPL
ChkStd - 1	ChkStd	24839	7.3861	-245.1		PPB	27 Jul 2018 14:54:37	U	SMPL
PB111510BL - 1	PBW	-164	-0.0415	-245.1		PPB	27 Jul 2018 14:56:46	U	SMPL
PB111510BS - 1	LCSW	12543	3.7334	-245.1		PPB	27 Jul 2018 14:58:55	U	SMPL
J4064-01 - 1	EFF-DAY-1-METAL	266	0.0863	-245.1		PPB	27 Jul 2018 15:05:53	U	SMPL
J4064-01DUP - 1	EFF-DAY-1-METALDUP	222	0.0732	-245.1		PPB	27 Jul 2018 15:08:00	U	SMPL
J4064-01MS - 1	EFF-DAY-1-METALMS	12710	3.7830	-245.1		PPB	27 Jul 2018 15:10:08	U	SMPL
J4064-01MSD - 1	EFF-DAY-1-METALMSD	12463	3.7096	-245.1		PPB	27 Jul 2018 15:12:17	U	SMPL
J4146-01 - 1	SES-CS-PH4-COA-01	-99	-0.0222	-245.1		PPB	27 Jul 2018 15:14:24	U	SMPL
CCV32 - 1	CCV32	15580	4.6355	-245.1		PPB	27 Jul 2018 15:16:31	U	SMPL
CCB32 - 1	CCB32	-188	-0.0486	-245.1		PPB	27 Jul 2018 15:18:38	U	SMPL
J4160-01 - 1	072418-TIPC-F-D-B	-13	0.0034	-245.1		PPB	27 Jul 2018 15:20:46	U	SMPL
J4160-02 - 1	072418-TIPC-F-D-S	9	0.0099	-245.1		PPB	27 Jul 2018 15:22:53	U	SMPL
J4160-03 - 1	072418-TIPC-F-U-B	-18	0.0019	-245.1		PPB	27 Jul 2018 15:25:00	U	SMPL
J4160-04 - 1	072418-TIPC-F-U-S	-5	0.0057	-245.1		PPB	27 Jul 2018 15:27:07	U	SMPL
J4196-01 - 1	072518-TIPC-E-D-M	-30	-0.0017	-245.1		PPB	27 Jul 2018 15:29:14	U	SMPL
J4196-02 - 1	072518-TIPC-E-U-M	-41	-0.0049	-245.1		PPB	27 Jul 2018 15:31:21	U	SMPL
J4196-03 - 1	072518-DBRI-E-D-B	-29	-0.0014	-245.1		PPB	27 Jul 2018 15:33:29	U	SMPL
J4196-04 - 1	072518-DBRI-E-D-M	0	0.0072	-245.1		PPB	27 Jul 2018 15:35:38	U	SMPL
J4196-05 - 1	072518-DBRI-E-D-S	-54	-0.0088	-245.1		PPB	27 Jul 2018 15:37:46	U	SMPL
J4196-06 - 1	072518-DBRI-E-U-B	-30	-0.0017	-245.1		PPB	27 Jul 2018 15:39:53	U	SMPL
CCV33 - 1	CCV33	15860	4.7187	-245.1		PPB	27 Jul 2018 15:42:00	U	SMPL
CCB33 - 1	CCB33	-201	-0.0525	-245.1		PPB	27 Jul 2018 15:44:07	U	SMPL
J4196-07 - 1	072518-DBRI-E-U-M	-61	-0.0109	-245.1		PPB	27 Jul 2018 15:46:15	U	SMPL
J4196-08 - 1	072518-DBRI-E-U-S	-61	-0.0109	-245.1		PPB	27 Jul 2018 15:48:22	U	SMPL
J4064-01LX5 - 1		-134	-0.0326	-245.1		PPB	27 Jul 2018 15:50:29	U	SMPL
J4064-01A - 1		12307	3.6632	-245.1		PPB	27 Jul 2018 15:52:36	U	SMPL
CCV34 - 1	CCV34	15637	4.6525	-245.1		PPB	27 Jul 2018 15:54:44	U	SMPL
CCB34 - 1	CCB34	-279	-0.0756	-245.1		PPB	27 Jul 2018 15:56:51	U	SMPL