



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96942

Slope : 98.9

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/27/2018	14:30
2	CAL2	1	Water			21.6	7.01	07/27/2018	14:32
3	CAL3	1	Water			21.8	10.04	07/27/2018	14:34
4	ICV	1	Water			21.7	7.00	07/27/2018	14:36
5	CCV1	1	Water			21.6	2.00	07/27/2018	14:38
6	J4216-01	1	Water			21.4	5.04	07/27/2018	14:40
7	J4216-01DUP	1	Water			21.5	5.05	07/27/2018	14:42
8	J4216-02	1	Water			21.5	4.62	07/27/2018	14:44
9	J4216-03	1	Water			21.6	6.30	07/27/2018	14:46
10	CCV2	1	Water			21.7	12.00	07/27/2018	14:48

WORKLIST(Hardcopy Internal Chain)

LB96942

WorkList Name : pH072718-1

WorkList ID : 115012

Date : 7/27/2018 2:15:37 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/26/2018	Water	J4216-01	A pH	Cool 4 deg C	STER09	N62	CELL-A-POOL-WATER(DE-IC	07/26/2018	SM 4500-PH B
07/26/2018	Water	J4216-02	A pH	Cool 4 deg C	STER09	N62	CELL-B-C-POOL-WATER(DE-	07/26/2018	SM 4500-PH B
07/26/2018	Water	J4216-03	A pH	Cool 4 deg C	STER09	N62	CELL-D-POOL-WATER(DE-IC	07/26/2018	SM 4500-PH B

Date/Time 07/27/18 2:20

Received by: SP

Relinquished by: CP

Date/Time 07/27/18 3:00

Received by: CP

Relinquished by: SP