

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB96943

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2337
IODINE SOLUTION .025N 1L	W2349
Starch Solution, 4L	W2277

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB111531BL		1	50	50	2.00	0.00	1.74	1.74	0.26	0.00	0.00	07/27/2018	13:20
2	J4176-02		1	50	50	2.00	0.00	1.42	1.42	0.58	0.32	2.56	07/27/2018	13:22
3	J4176-02DUP		1	50	50	2.00	1.42	2.82	1.40	0.60	0.34	2.72	07/27/2018	13:25
4	J4188-02		1	50	50	2.00	2.82	4.02	1.20	0.80	0.54	4.32	07/27/2018	13:28
5	J4189-27		1	50	50	2.00	4.02	5.38	1.36	0.64	0.38	3.04	07/27/2018	13:30

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume