



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96948

Slope : 98.9

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/27/2018	14:30
2	CAL2	1	Water			21.6	7.01	07/27/2018	14:32
3	CAL3	1	Water			21.8	10.04	07/27/2018	14:34
4	ICV	1	Water			21.7	7.00	07/27/2018	14:36
5	CCV1	1	Water			21.7	2.00	07/27/2018	15:10
6	J4200-05	1	Water			27.8	6.65	07/27/2018	15:12
7	J4200-05DUP	1	Water			27.9	6.66	07/27/2018	15:14
8	CCV2	1	Water			21.6	12.00	07/27/2018	15:16

LB96948

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH072718-2

WorkList ID : 115016

Date : 7/27/2018 2:52:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/27/2018	Water	J4200-05	pH	Cool 4 deg C	PSEG01	K63	CA-DRUM-2-BASELINEWATE	07/27/2018	9040C

Date/Time

07/27/18 3:00

Received by:

SP

Relinquished by:

CP

Date/Time

07/27/18 3:30

Received by:

CP

Relinquished by:

SP