

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/30/2018 9:13

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.016	0.0	0.114	
ICB1	0.002	0.0	0.005	
CCV1	0.959	0.0	0.108	
CCB1	-0.001	0.0	0.004	
PB111506BL	0.000	0.0	0.005	
PB111506BS	1.017	0.0	0.114	
J4146-01	-0.001	0.0	0.004	
J4146-01DUP	0.000	0.0	0.005	
J4146-01MS	0.990	0.0	0.111	
J4146-01MSD	1.008	0.0	0.113	
CCV2	0.976	0.0	0.109	
CCB2	-0.002	0.0	0.004	

N	12
Mean	0.497
SD	0.5195
CV%	104.54

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 30 08:51:17 2018

Mon Jul 30 09:13:32 2018

Sample Id	Sam/Ctr/c	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0494	mg/l	7/30/2018 8:51:17	
0.1PPM	A	Phenolics-	P	0.1003	mg/l	7/30/2018 8:51:18	
0.25PPM	A	Phenolics-	P	0.2497	mg/l	7/30/2018 8:51:19	
0.50PPM	A	Phenolics-	P	0.5048	mg/l	7/30/2018 8:51:20	
1.0PPM	A	Phenolics-	P	0.9938	mg/l	7/30/2018 8:51:21	
2.0PPM	A	Phenolics-	P	2.0019	mg/l	7/30/2018 8:51:22	
ICV1	S	Phenolics-	P	1.0156	mg/l	7/30/2018 9:13:21	
ICB1	S	Phenolics-	P	0.0022	mg/l	7/30/2018 9:13:22	
CCV1	S	Phenolics-	P	0.9586	mg/l	7/30/2018 9:13:23	
CCB1	S	Phenolics-	P	-0.0006	mg/l	7/30/2018 9:13:24	
PB111506BL	S	Phenolics-	P	0.0002	mg/l	7/30/2018 9:13:25	
PB111506BS	S	Phenolics-	P	1.0168	mg/l	7/30/2018 9:13:26	
J4146-01	S	Phenolics-	P	-0.0007	mg/l	7/30/2018 9:13:27	
J4146-01DUP	S	Phenolics-	P	0	mg/l	7/30/2018 9:13:28	
J4146-01MS	S	Phenolics-	P	0.9896	mg/l	7/30/2018 9:13:29	
J4146-01MSD	S	Phenolics-	P	1.0081	mg/l	7/30/2018 9:13:30	
CCV2	S	Phenolics-	P	0.9756	mg/l	7/30/2018 9:13:31	
CCB2	S	Phenolics-	P	-0.0025	mg/l	7/30/2018 9:13:32	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

7/30/2018 8:51


Test Phenolics-

Accepted 7/30/2018 8:51

Factor 9.303
 Bias 0.005

Coeff. of det. 0.999976

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors

1	Phenol-2	0.010	0.0494	0.0500	
2	Phenol-2	0.015	0.1003	0.1000	
3	Phenol-2	0.031	0.2497	0.2500	
4	Phenol-2	0.059	0.5048	0.5000	
5	Phenol-2	0.111	0.9938	1.0000	
6	Phenol-2	0.220	2.0019	2.0000	