



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96965

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.6	07/30/2018	09:20
2	J4200-06		1	7	>100.0	>212.0	>212.0	07/30/2018	09:45
3	J4200-06DUP		1	7	>100.0	>212.0	>212.0	07/30/2018	10:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WorkList Name : FP73018

WorkList ID : 115044

Date : 7/30/2018 8:39:47 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/01/2018	Water	J4200-06	C Flash Point	Cool 4 deg C	PSEG01	K63	CA-DRUM-2-BASELINEWATE	07/27/2018	1010A

Date/Time 07/30/18 08:45 AM

Received by: KS

Relinquished by: CP

Date/Time 07/30/18 10:30 AM

Received by: CP

Relinquished by: KS

WORKLIST(Hardcopy Internal Chain)

LB96965