


7470A

LB 96966
INSTRUMENT ID: CV2

Linear

 μ Abs.:

37084



A= 0.0000e+000

B= 2.7041e-004

C= 4.0074e-002

Rho= 0.9998545

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.048	0.048	30	0.000	30				
0.2	0.200	0.236	0.036	723	0.0 %	723				
2.5	2.500	2.476	-0.024	9007	0.0 %	9007				
5.0	5.000	4.878	-0.122	17890	0.0 %	17890				
7.5	7.500	7.495	-0.005	27570	0.0 %	27570				
10.0	10.000	10.068	0.068	37084	0.0 %	37084				

LB96966 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Jul 2018 10:34:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	30	-	0.0000	7470A	PPB	30 Jul 2018 10:52:32	S	Std
0.2 - 1	50.2	723	-	0.2000	7470A	PPB	30 Jul 2018 10:54:40	S	Std
2.5 - 1	52.5	9007	-	2.5000	7470A	PPB	30 Jul 2018 10:56:48	S	Std
5.0 - 1	55	17890	-	5.0000	7470A	PPB	30 Jul 2018 10:58:56	S	Std
7.5 - 1	57.5	27570	-	7.5000	7470A	PPB	30 Jul 2018 11:01:04	S	Std
10.0 - 1	510	37084	-	10.0000	7470A	PPB	30 Jul 2018 11:03:12	S	Std
ICV77 - 1	ICV77	13992	3.8236	-7470A	PPB	PPB	30 Jul 2018 11:05:50	U	SMPL
ICB77 - 1	ICB77	-166	-0.0048	-7470A	PPB	PPB	30 Jul 2018 11:07:56	U	SMPL
CCV39 - 1	CCV39	18365	5.0061	-7470A	PPB	PPB	30 Jul 2018 11:10:04	U	SMPL
CCB39 - 1	CCB39	-173	-0.0067	-7470A	PPB	PPB	30 Jul 2018 11:12:10	U	SMPL
CRA - 1	CRA	706	0.2310	-7470A	PPB	PPB	30 Jul 2018 11:14:18	U	SMPL
HighStd - 1	HighStd	38071	10.3347	-7470A	PPB	PPB	30 Jul 2018 11:16:24	U	SMPL
ChkStd - 1	ChkStd	24878	6.7672	-7470A	PPB	PPB	30 Jul 2018 11:18:31	U	SMPL
PB111561BL - 1	PBW	-251	-0.0278	-7470A	PPB	PPB	30 Jul 2018 11:20:40	U	SMPL
PB111561BS - 1	LCSW	13656	3.7327	-7470A	PPB	PPB	30 Jul 2018 11:22:50	U	SMPL
PB111561TB - 1	PB111561TB PB111512TB	-146	0.0006	-7470A	PPB	PPB	30 Jul 2018 11:24:57	U	SMPL
J4176-05 - 1	PLASTIC-COAT	13	0.0436	-7470A	PPB	PPB	30 Jul 2018 11:27:05	U	SMPL
J4177-01 - 1	SITE1-IDW-COMP-072518	8	0.0422	-7470A	PPB	PPB	30 Jul 2018 11:29:14	U	SMPL
J4180-04 - 1	GS-PIT-1	33	0.0490	-7470A	PPB	PPB	30 Jul 2018 11:31:22	U	SMPL
J4180-08 - 1	GS-PIT-2	0	0.0401	-7470A	PPB	PPB	30 Jul 2018 11:33:29	U	SMPL
CCV40 - 1	CCV40	18206	4.9631	-7470A	PPB	PPB	30 Jul 2018 11:35:36	U	SMPL
CCB40 - 1	CCB40	-180	-0.0086	-7470A	PPB	PPB	30 Jul 2018 11:37:43	U	SMPL
J4189-04 - 1	TP-12-S	-21	0.0344	-7470A	PPB	PPB	30 Jul 2018 11:39:50	U	SMPL
J4189-08 - 1	TP-12-D	-2	0.0395	-7470A	PPB	PPB	30 Jul 2018 11:41:57	U	SMPL
J4189-12 - 1	TP-13-S	11	0.0430	-7470A	PPB	PPB	30 Jul 2018 11:44:04	U	SMPL
J4189-16 - 1	TP-14-S	7	0.0420	-7470A	PPB	PPB	30 Jul 2018 11:46:10	U	SMPL
J4189-20 - 1	TP-14-D	3	0.0409	-7470A	PPB	PPB	30 Jul 2018 11:48:18	U	SMPL
J4189-20DUP - 1	TP-14-DDUP	-5	0.0387	-7470A	PPB	PPB	30 Jul 2018 11:50:25	U	SMPL
J4189-20MS - 1	TP-14-DMS	14079	3.8471	-7470A	PPB	PPB	30 Jul 2018 11:52:33	U	SMPL
J4189-20MSD - 1	TP-14-DMSD	14075	3.8460	-7470A	PPB	PPB	30 Jul 2018 11:54:42	U	SMPL
PB111562BL - 1	PBW	-130	0.0049	-7470A	PPB	PPB	30 Jul 2018 11:56:50	U	SMPL
PB111562BS - 1	LCSW	13664	3.7349	-7470A	PPB	PPB	30 Jul 2018 11:58:57	U	SMPL
CCV41 - 1	CCV41	18438	5.0258	-7470A	PPB	PPB	30 Jul 2018 12:02:43	U	SMPL
CCB41 - 1	CCB41	-167	-0.0051	-7470A	PPB	PPB	30 Jul 2018 12:04:50	U	SMPL
PB111562TB - 1	PB111514TB	-11	0.0371	-7470A	PPB	PPB	30 Jul 2018 12:06:58	U	SMPL
J4176-02 - 1	SV26835	271	0.1134	-7470A	PPB	PPB	30 Jul 2018 12:09:05	U	SMPL
J4176-02DUP - 1	SV26835DUP	257	0.1096	-7470A	PPB	PPB	30 Jul 2018 12:11:12	U	SMPL
J4176-02MS - 1	SV26835MS	13882	3.7939	-7470A	PPB	PPB	30 Jul 2018 12:13:19	U	SMPL
J4176-02MSD - 1	SV26835MSD	14358	3.9226	-7470A	PPB	PPB	30 Jul 2018 12:15:26	U	SMPL
J4188-02 - 1	HA-BASELINEWATER-N48971	-138	0.0028	-7470A	PPB	PPB	30 Jul 2018 12:17:35	U	SMPL
J4189-27 - 1	TP-12	7	0.0420	-7470A	PPB	PPB	30 Jul 2018 12:19:43	U	SMPL
CCV42 - 1	CCV42	18224	4.9680	-7470A	PPB	PPB	30 Jul 2018 12:21:51	U	SMPL
CCB42 - 1	CCB42	-172	-0.0064	-7470A	PPB	PPB	30 Jul 2018 12:24:00	U	SMPL
J4189-20LX5 - 1		-9	0.0376	-7470A	PPB	PPB	30 Jul 2018 12:26:08	U	SMPL
J4189-20A - 1		15246	4.1627	-7470A	PPB	PPB	30 Jul 2018 12:28:15	U	SMPL
J4176-02LX5 - 1		-93	0.0149	-7470A	PPB	PPB	30 Jul 2018 12:30:22	U	SMPL
J4176-02A - 1		15400	4.2043	-7470A	PPB	PPB	30 Jul 2018 12:32:29	U	SMPL
CCV43 - 1	CCV43	18147	4.9471	-7470A	PPB	PPB	30 Jul 2018 12:34:37	U	SMPL
CCB43 - 1	CCB43	-201	-0.0143	-7470A	PPB	PPB	30 Jul 2018 12:36:45	U	SMPL