


7471B

L396980
INSTRUMENT ID: CV1

Linear

 μ Abs.:

48751



A= 0.0000e+000

B= 2.0544e-004

C= -1.5600e-002

Rho= 0.9999973

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	58	0.000	58				
0.2	0.200	0.198	-0.002	1038	0.0 %	1038				
2.5	2.500	2.500	0.000	12247	0.0 %	12247				
5.0	5.000	5.017	0.017	24497	0.0 %	24497				
7.5	7.500	7.489	-0.011	36529	0.0 %	36529				
10.0	10.000	10.000	-0.000	48751	0.0 %	48751				

LB96980 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Jul 2018 16:07:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	58	-	0.0000	7471B	PPB	30 Jul 2018 16:10:53	S	Std
0.2 - 1	50.2	1038	-	0.2000	7471B	PPB	30 Jul 2018 16:13:11	S	Std
2.5 - 1	52.5	12247	-	2.5000	7471B	PPB	30 Jul 2018 16:15:29	S	Std
5.0 - 1	55	24497	-	5.0000	7471B	PPB	30 Jul 2018 16:17:46	S	Std
7.5 - 1	57.5	36529	-	7.5000	7471B	PPB	30 Jul 2018 16:20:03	S	Std
10.0 - 1	510	48751	-	10.0000	7471B	PPB	30 Jul 2018 16:22:20	S	Std
ICV78 - 1	ICV78	19257	3.9405	-	7471B	PPB	30 Jul 2018 16:28:51	U	SMPL
ICB78 - 1	ICB78	-157	-0.0479	-	7471B	PPB	30 Jul 2018 16:31:06	U	SMPL
CCV46 - 1	CCV46	24529	5.0236	-	7471B	PPB	30 Jul 2018 16:33:23	U	SMPL
CCB46 - 1	CCB46	-181	-0.0528	-	7471B	PPB	30 Jul 2018 16:35:38	U	SMPL
CRA - 1	CRA	1002	0.1902	-	7471B	PPB	30 Jul 2018 16:37:55	U	SMPL
HighStd - 1	HighStd	48987	10.0482	-	7471B	PPB	30 Jul 2018 16:40:11	U	SMPL
ChkStd - 1	ChkStd	33870	6.9426	-	7471B	PPB	30 Jul 2018 16:42:27	U	SMPL
PB111572BL - 1	PBS	-273	-0.0717	-	7471B	PPB	30 Jul 2018 16:44:46	U	SMPL
PB111572BS - 1	LCSS	16748	3.4251	-	7471B	PPB	30 Jul 2018 16:47:05	U	SMPL
J3831-03 - 1	SU-03-072718	3429	0.6888	-	7471B	PPB	30 Jul 2018 16:49:21	U	SMPL
J4210-01 - 1	TOPSOIL-072618	4773	0.9650	-	7471B	PPB	30 Jul 2018 16:51:39	U	SMPL
J4217-01 - 1	SOIL-1	1924	0.3797	-	7471B	PPB	30 Jul 2018 16:53:57	U	SMPL
J4217-01DUP - 1	SOIL-1DUP	2164	0.4290	-	7471B	PPB	30 Jul 2018 16:56:15	U	SMPL
J4217-01MS - 1	SOIL-1MS	19930	4.0788	-	7471B	PPB	30 Jul 2018 16:58:31	U	SMPL
CCV47 - 1	CCV47	24105	4.9365	-	7471B	PPB	30 Jul 2018 17:00:46	U	SMPL
CCB47 - 1	CCB47	-189	-0.0544	-	7471B	PPB	30 Jul 2018 17:03:03	U	SMPL
J4217-01MSD - 1	SOIL-1MSD	18959	3.8793	-	7471B	PPB	30 Jul 2018 17:05:20	U	SMPL
J4217-05 - 1	SOIL-2	4345	0.8770	-	7471B	PPB	30 Jul 2018 17:07:36	U	SMPL
J4218-03 - 1	TP-15-S	3934	0.7926	-	7471B	PPB	30 Jul 2018 17:09:52	U	SMPL
J4218-07 - 1	TP-15-D	896	0.1685	-	7471B	PPB	30 Jul 2018 17:12:08	U	SMPL
J4217-01LX5 - 1		336	0.0534	-	7471B	PPB	30 Jul 2018 17:14:24	U	SMPL
J4217-01A - 1		20188	4.1318	-	7471B	PPB	30 Jul 2018 17:16:42	U	SMPL
CCV48 - 1	CCV48	24173	4.9505	-	7471B	PPB	30 Jul 2018 17:18:58	U	SMPL
CCB48 - 1	CCB48	-181	-0.0528	-	7471B	PPB	30 Jul 2018 17:21:16	U	SMPL