



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB96982
Analysis Date: 07/30/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 07/30/2018
Extraction IN Time: 07:40
Extraction OUT Time: 08:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB96982BL	LB96982BL	WATER	2	1000	100	3.1844	3.1844	3.02	3.1844	3.1844	0.0000	0
2	LB96982BS	LB96982BS	WATER	2	1000	100	3.0266	3.0266	3.01	3.0454	3.0454	0.0188	18.8
3	LB96982BSD	LB96982BSD	WATER	2	1000	100	2.1847	2.1847	3.03	2.2037	2.2037	0.0190	19
4	J4200-05	CA-DRUM-2-BASELINEWATER	WATER	2	1000	100	3.2220	3.2220	3.04	3.2643	3.2643	0.0423	42.3



QC Batch# LB96982 Test: TPH Analysis Date: 07/30/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2392
pH Paper 0-14	W2329
Sodium Sulfate	EP1866
1:1 HCL	WP64228
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	10.00 ML	WP64230
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>09:46</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>08:40</u>		
Bal Check Time:	<u>08:00</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:15</u>
		Out Time:	<u>09:45</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:16</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>10:40</u>		
Bal Check Time:	<u>12:10</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:40</u>
		Out Time:	<u>11:15</u>		

2846463
18 2896982

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph j4200

WorkList ID : 115024

Date : 7/30/2018 7:30:10 am

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/01/2018	Water	J4200-05 -K	TPH	Conc H2SO4 to	PSEG01	K63	CA-DRUM-2-BASELINEWATE	07/27/2018	1664A

Date/Time 07/30/18 7:35 AM
Received by: JN
Relinquished by: CB

Date/Time 07/30/18 10:30 AM
Received by: CJ
Relinquished by: JP