

LB 96988


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37692



A= 0.0000e+000

B= 2.6713e-004

C= 2.5125e-002

Rho= 0.9994446

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	-58	0.000	-58				
0.2	0.200	0.202	0.002	661	0.0 %	661				
2.5	2.500	2.468	-0.032	9143	0.0 %	9143				
5.0	5.000	5.161	0.161	19227	0.0 %	19227				
7.5	7.500	7.266	-0.234	27106	0.0 %	27106				
10.0	10.000	10.094	0.094	37692	0.0 %	37692				

LB96988 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Jul 2018 10:42:26

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-58	-	0.0007470A	PPB		31 Jul 2018 10:51:40	S	Std
0.2 - 1	50.2	661	-	0.20007470A	PPB		31 Jul 2018 10:53:49	S	Std
2.5 - 1	52.5	9143	-	2.50007470A	PPB		31 Jul 2018 10:55:58	S	Std
5.0 - 1	55	19227	-	5.00007470A	PPB		31 Jul 2018 10:58:06	S	Std
7.5 - 1	57.5	27106	-	7.50007470A	PPB		31 Jul 2018 11:00:14	S	Std
10.0 - 1	510	37692	-	10.00007470A	PPB		31 Jul 2018 11:02:22	S	Std
ICV79 - 1	ICV79	14741	3.9629	-7470A	PPB		31 Jul 2018 11:05:00	U	SMPL
ICB79 - 1	ICB79	-262	-0.0449	-7470A	PPB		31 Jul 2018 11:07:06	U	SMPL
CCV49 - 1	CCV49	19196	5.1530	-7470A	PPB		31 Jul 2018 11:09:14	U	SMPL
CCB49 - 1	CCB49	-276	-0.0486	-7470A	PPB		31 Jul 2018 11:11:20	U	SMPL
CRA - 1	CRA	613	0.1889	-7470A	PPB		31 Jul 2018 11:13:29	U	SMPL
HighStd - 1	HighStd	38601	10.3367	-7470A	PPB		31 Jul 2018 11:15:35	U	SMPL
ChkStd - 1	ChkStd	26610	7.1335	-7470A	PPB		31 Jul 2018 11:20:13	U	SMPL
PB111602BL - 1	PBW	-362	-0.0716	-7470A	PPB		31 Jul 2018 11:22:21	U	SMPL
PB111602BS - 1	LCSW	13689	3.6819	-7470A	PPB		31 Jul 2018 11:24:30	U	SMPL
PB111602TB - 1	PB111569TB	-209	-0.0307	-7470A	PPB		31 Jul 2018 11:26:38	U	SMPL
J3831-04 - 1	SU-03-072718	-108	-0.0037	-7470A	PPB		31 Jul 2018 11:28:47	U	SMPL
J3831-04DUP - 1	SU-03-072718DUP	-88	0.0016	-7470A	PPB		31 Jul 2018 11:30:55	U	SMPL
J3831-04MS - 1	SU-03-072718MS	14046	3.7773	-7470A	PPB		31 Jul 2018 11:33:04	U	SMPL
J3831-04MSD - 1	SU-03-072718MSD	14664	3.9424	-7470A	PPB		31 Jul 2018 11:35:11	U	SMPL
CCV50 - 1	CCV50	19296	5.1797	-7470A	PPB		31 Jul 2018 11:37:18	U	SMPL
CCB50 - 1	CCB50	-275	-0.0483	-7470A	PPB		31 Jul 2018 11:39:26	U	SMPL
J4200-01 - 1	CA-DRUM-1-WIPES	-133	-0.0104	-7470A	PPB		31 Jul 2018 11:41:34	U	SMPL
J4200-02 - 1	CA-DRUM-1-SPEEDY-DRY	-90	0.0011	-7470A	PPB		31 Jul 2018 11:43:41	U	SMPL
J4213-01 - 1	SS-WIPE-DRUM-1	-69	0.0067	-7470A	PPB		31 Jul 2018 11:45:48	U	SMPL
J4217-04 - 1	SOIL-1	-85	0.0024	-7470A	PPB		31 Jul 2018 11:47:55	U	SMPL
J4217-08 - 1	SOIL-2	-69	0.0067	-7470A	PPB		31 Jul 2018 11:50:02	U	SMPL
J4218-04 - 1	TP-15-S	-30	0.0171	-7470A	PPB		31 Jul 2018 11:52:10	U	SMPL
J4218-08 - 1	TP-15-D	-72	0.0059	-7470A	PPB		31 Jul 2018 11:54:18	U	SMPL
PB111603BL - 1	PBW	-71	0.0062	-7470A	PPB		31 Jul 2018 11:56:26	U	SMPL
PB111603BS - 1	LCSW	13594	3.6565	-7470A	PPB		31 Jul 2018 11:58:34	U	SMPL
PB111603TB - 1	PB111571TB	-206	-0.0299	-7470A	PPB		31 Jul 2018 12:00:41	U	SMPL
CCV51 - 1	CCV51	19490	5.2315	-7470A	PPB		31 Jul 2018 12:02:49	U	SMPL
CCB51 - 1	CCB51	-263	-0.0451	-7470A	PPB		31 Jul 2018 12:04:56	U	SMPL
J4200-06 - 1	CA-DRUM-2-BASELINEWATER	-113	-0.0051	-7470A	PPB		31 Jul 2018 12:07:04	U	SMPL
J4200-06DUP - 1	CA-DRUM-2-BASELINEWATERDUP	-77	0.0046	-7470A	PPB		31 Jul 2018 12:09:11	U	SMPL
J4200-06MS - 1	CA-DRUM-2-BASELINEWATERMS	14067	3.7829	-7470A	PPB		31 Jul 2018 12:11:19	U	SMPL
J4200-06MSD - 1	CA-DRUM-2-BASELINEWATERMSD	13551	3.6450	-7470A	PPB		31 Jul 2018 12:13:26	U	SMPL
PB111604BL - 1	PBW	-261	-0.0446	-7470A	PPB		31 Jul 2018 12:15:34	U	SMPL
PB111604BS - 1	LCSW	13807	3.7134	-7470A	PPB		31 Jul 2018 12:17:43	U	SMPL
J4200-05 - 1	CA-DRUM-2-BASELINEWATER	-91	0.0008	-7470A	PPB		31 Jul 2018 12:19:50	U	SMPL
J4200-05DUP - 1	CA-DRUM-2-BASELINEWATERDUP	-30	0.0171	-7470A	PPB		31 Jul 2018 12:21:59	U	SMPL
J4200-05MS - 1	CA-DRUM-2-BASELINEWATERMS	13459	3.6205	-7470A	PPB		31 Jul 2018 12:26:56	U	SMPL
J4200-05MSD - 1	CA-DRUM-2-BASELINEWATERMSD	14192	3.8163	-7470A	PPB		31 Jul 2018 12:29:03	U	SMPL
CCV52 - 1	CCV52	19082	5.1225	-7470A	PPB		31 Jul 2018 12:31:10	U	SMPL
CCB52 - 1	CCB52	-300	-0.0550	-7470A	PPB		31 Jul 2018 12:33:18	U	SMPL
J3831-04LX5 - 1		-120	-0.0069	-7470A	PPB		31 Jul 2018 12:35:27	U	SMPL
J3831-04A - 1		14234	3.8275	-7470A	PPB		31 Jul 2018 12:37:34	U	SMPL
J4200-06LX5 - 1		-255	-0.0430	-7470A	PPB		31 Jul 2018 12:39:41	U	SMPL
J4200-06A - 1		14135	3.8010	-7470A	PPB		31 Jul 2018 12:41:49	U	SMPL
J4200-05LX5 - 1		-99	-0.0013	-7470A	PPB		31 Jul 2018 13:01:17	U	SMPL
J4200-05A - 1		12282	3.3060	-7470A	PPB		31 Jul 2018 13:03:25	U	SMPL
CCV53 - 1	CCV53	19354	5.1952	-7470A	PPB		31 Jul 2018 13:05:33	U	SMPL
CCB53 - 1	CCB53	-299	-0.0547	-7470A	PPB		31 Jul 2018 13:07:42	U	SMPL