



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96989

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP66572
Turbidity std, 40NTU	WP66566
turbidity std 5 NTU	WP66569
turbidity std 20 NTU	WP66567
Turbidity std 1 NTU	WP66570
turbidity std 0 NTU	WP66571
turbidity std 10 NTU	WP66568

Intercept: -0.0879

Slope: 0.9461

Regression: 0.9977

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.269	07/31/2018	09:52
2	CAL2	1	1	0.996	07/31/2018	09:56
3	CAL3	5	1	4.860	07/31/2018	09:59
4	CAL4	10	1	9.760	07/31/2018	10:03
5	CAL5	20	1	16.890	07/31/2018	10:06
6	CAL6	40	1	38.600	07/31/2018	10:11



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96989

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	8.760	9.352	07/31/2018	10:14
2	ICB		1	0.260	0.368	07/31/2018	10:16
3	CCV1	10	1	9.450	10.081	07/31/2018	10:20
4	CCB1		1	0.261	0.369	07/31/2018	10:24
5	LB96989BL		1	0.259	0.367	07/31/2018	10:27
6	J4229-01		1	0.836	0.977	07/31/2018	10:30
7	J4229-01DUP		1	0.842	0.983	07/31/2018	10:33
8	CCV2	10	1	9.400	10.028	07/31/2018	10:36
9	CCB2		1	0.252	0.359	07/31/2018	10:39

WORKLIST(Hardcopy Internal Chain)

CB96989

WorkList Name : turbidity073118

WorkList ID : 115097

Date : 7/31/2018 9:03:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/31/2018	Water	J4229-01	Turbidity	1:1 HNO3 to p	HOME01	--Select--	BLUEBERRY-LAKE-PW	07/29/2018	SM2130 B

Date/Time 7-31-18 9:10
 Received by: AR
 Relinquished by: CP

Date/Time 7-31-18 10:45
 Received by: CP
 Relinquished by: AR