

245.1


LB97003

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35015



A= 0.0000e+000

B= 2.8683e-004

C= -2.4076e-002

Rho= 0.9999116

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.036	-0.036	-42	0.000	-42				
0.05	0.050	0.040	-0.010	225	0.0 %	225				
0.2	0.200	0.192	-0.008	753	0.0 %	753				
2.5	2.500	2.564	0.064	9023	0.0 %	9023				
5.0	5.000	5.058	0.058	17720	0.0 %	17720				
7.5	7.500	7.412	-0.088	25926	0.0 %	25926				
10.0	10.000	10.019	0.019	35015	0.0 %	35015				

LB97003 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 31 Jul 2018 13:57:52				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-42	-	0.0000245.1	PPB	31 Jul 2018 14:15:42	S	Std
0.05 - 1	50.05	225	-	0.0500245.1	PPB	31 Jul 2018 14:17:50	S	Std
0.2 - 1	50.2	753	-	0.2000245.1	PPB	31 Jul 2018 14:19:58	S	Std
2.5 - 1	52.5	9023	-	2.5000245.1	PPB	31 Jul 2018 14:22:06	S	Std
5.0 - 1	55	17720	-	5.0000245.1	PPB	31 Jul 2018 14:24:14	S	Std
7.5 - 1	57.5	25926	-	7.5000245.1	PPB	31 Jul 2018 14:26:21	S	Std
10.0 - 1	510	35015	-	10.0000245.1	PPB	31 Jul 2018 14:28:29	S	Std
ICV80 - 1	ICV80	13627	3.8845	-245.1	PPB	31 Jul 2018 14:31:10	U	SMPL
ICB80 - 1	ICB80	-60	-0.0413	-245.1	PPB	31 Jul 2018 14:33:17	U	SMPL
CCV54 - 1	CCV54	17845	5.0943	-245.1	PPB	31 Jul 2018 14:35:25	U	SMPL
CCB54 - 1	CCB54	-99	-0.0525	-245.1	PPB	31 Jul 2018 14:37:31	U	SMPL
CRA - 1	CRA	270	0.0534	-245.1	PPB	31 Jul 2018 14:42:05	U	SMPL
HighStd - 1	HighStd	34998	10.0143	-245.1	PPB	31 Jul 2018 14:44:11	U	SMPL
ChkStd - 1	ChkStd	24742	7.0726	-245.1	PPB	31 Jul 2018 14:46:17	U	SMPL
PB111608BL - 1	PBW	-18	-0.0292	-245.1	PPB	31 Jul 2018 14:50:54	U	SMPL
PB111608BS - 1	LCSW	13553	3.8633	-245.1	PPB	31 Jul 2018 14:53:01	U	SMPL
J4045-05 - 1	PT-HG-WS	23802	6.8030	-245.1	PPB	31 Jul 2018 14:55:08	U	SMPL
J4219-01 - 1	072618-DBRI-F-D-B	-83	-0.0479	-245.1	PPB	31 Jul 2018 14:59:45	U	SMPL
J4219-02 - 1	072618-DBRI-F-D-M	-122	-0.0591	-245.1	PPB	31 Jul 2018 15:01:53	U	SMPL
J4219-03 - 1	072618-DBRI-F-D-S	-102	-0.0533	-245.1	PPB	31 Jul 2018 15:04:01	U	SMPL
J4219-04 - 1	072618-DBRI-F-U-B	-101	-0.0530	-245.1	PPB	31 Jul 2018 15:06:08	U	SMPL
CCV55 - 1	CCV55	17645	5.0370	-245.1	PPB	31 Jul 2018 15:08:14	U	SMPL
CCB55 - 1	CCB55	-82	-0.0476	-245.1	PPB	31 Jul 2018 15:10:21	U	SMPL
J4219-05 - 1	072618-DBRI-F-U-M	-170	-0.0728	-245.1	PPB	31 Jul 2018 15:12:29	U	SMPL
J4219-06 - 1	072618-DBRI-F-U-S	-129	-0.0611	-245.1	PPB	31 Jul 2018 15:14:36	U	SMPL
J4219-07 - 1	072618-TIPC-F-D-B	-109	-0.0553	-245.1	PPB	31 Jul 2018 15:16:43	U	SMPL
J4219-08 - 1	072618-TIPC-F-D-S	-79	-0.0467	-245.1	PPB	31 Jul 2018 15:18:50	U	SMPL
J4219-09 - 1	072618-TIPC-F-U-B	-89	-0.0496	-245.1	PPB	31 Jul 2018 15:20:58	U	SMPL
J4219-10 - 1	072618-TIPC-F-U-S	-92	-0.0505	-245.1	PPB	31 Jul 2018 15:23:06	U	SMPL
J4219-10DUP - 1	072618-TIPC-F-U-SDUP	-92	-0.0505	-245.1	PPB	31 Jul 2018 15:25:15	U	SMPL
J4219-10MS - 1	072618-TIPC-F-U-SMS	14061	4.0090	-245.1	PPB	31 Jul 2018 15:27:24	U	SMPL
J4219-10MSD - 1	072618-TIPC-F-U-SMSD	14430	4.1148	-245.1	PPB	31 Jul 2018 15:29:32	U	SMPL
J4229-01 - 1	BLUEBERRY-LAKE-PW	-225	-0.0886	-245.1	PPB	31 Jul 2018 15:31:39	U	SMPL
CCV56 - 1	CCV56	17532	5.0046	-245.1	PPB	31 Jul 2018 15:33:46	U	SMPL
CCB56 - 1	CCB56	-114	-0.0568	-245.1	PPB	31 Jul 2018 15:35:53	U	SMPL
J4219-10LX5 - 1		20	-0.0183	-245.1	PPB	31 Jul 2018 15:38:02	U	SMPL
J4219-10A - 1		16606	4.7390	-245.1	PPB	31 Jul 2018 15:40:09	U	SMPL
CCV57 - 1	CCV57	17651	5.0387	-245.1	PPB	31 Jul 2018 15:50:52	U	SMPL
CCB57 - 1	CCB57	-152	-0.0677	-245.1	PPB	31 Jul 2018 15:53:02	U	SMPL