

# TitriSoft Report

| Ident       | Amount | DateTime                    | Alkalinity       | Final pH (sample) | Initial pH (sample) | slope            | offset         |
|-------------|--------|-----------------------------|------------------|-------------------|---------------------|------------------|----------------|
| LB97004BLW  | 100    | <del>31/07/2018 15:06</del> | 0.00             | 3.54              | 3.55                | 104.6            | 7.1            |
| LB97004BSW  | 100    | 31/07/2018 15:12            | 46.76            | 4.30              | 8.37                | 104.6            | 7.1            |
| J4080-22    | 50     | 31/07/2018 15:23            | 122.24           | 4.43              | 6.97                | 104.6            | 7.1            |
| J4229-01    | 50     | 31/07/2018 15:32            | 82.08            | 4.43              | 5.89                | 104.6            | 7.1            |
| J4233-01    | 50     | 31/07/2018 15:36            | 28.48            | 4.33              | 5.65                | 104.6            | 7.1            |
| J4233-01DUP | 50     | 31/07/2018 15:40            | 29.04            | 4.33              | 5.61                | 104.6            | 7.1            |
| J4233-02    | 50     | 31/07/2018 15:43            | 21.12            | 4.27              | 5.59                | 104.6            | 7.1            |
| J4233-03    | 50     | <del>31/07/2018 15:47</del> | <del>18.24</del> | <del>4.22</del>   | <del>5.80</del>     | <del>104.6</del> | <del>7.1</del> |
| J4233-03    | 100    | 31/07/2018 16:40            | 16.40            | 4.36              | 5.85                | 104.6            | 7.1            |

07/31/18 IB