


7471B

LB97008
INSTRUMENT ID: CV2

Linear

 μ Abs.:

45486



A= 0.0000e+000

B= 2.2004e-004

C= 1.5680e-003

Rho= 0.9998350

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	25	0.000	25				
0.2	0.200	0.198	-0.002	892	0.0 %	892				
2.5	2.500	2.439	-0.061	11078	0.0 %	11078				
5.0	5.000	5.128	0.128	23297	0.0 %	23297				
7.5	7.500	7.418	-0.082	33704	0.0 %	33704				
10.0	10.000	10.010	0.010	45486	0.0 %	45486				

LB97008 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 Jul 2018 16:12:24

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.00007471B	PPB		31 Jul 2018 16:15:21	S	Std
0.2 - 1	50.2	892	-	0.20007471B	PPB		31 Jul 2018 16:17:39	S	Std
2.5 - 1	52.5	11078	-	2.50007471B	PPB		31 Jul 2018 16:19:57	S	Std
5.0 - 1	55	23297	-	5.00007471B	PPB		31 Jul 2018 16:22:14	S	Std
7.5 - 1	57.5	33704	-	7.50007471B	PPB		31 Jul 2018 16:24:31	S	Std
10.0 - 1	58	45486	-	10.00007471B	PPB		31 Jul 2018 16:29:28	S	Std
ICV81 - 1	ICV81	17717	3.9000	-7471B	PPB		31 Jul 2018 16:32:10	U	SMPL
ICB81 - 1	ICB81	-226	-0.0482	-7471B	PPB		31 Jul 2018 16:34:27	U	SMPL
CCV58 - 1	CCV58	21319	4.6926	-7471B	PPB		31 Jul 2018 16:36:43	U	SMPL
CCB58 - 1	CCB58	-219	-0.0466	-7471B	PPB		31 Jul 2018 16:38:59	U	SMPL
CRA - 1	CRA	851	0.1888	-7471B	PPB		31 Jul 2018 16:41:17	U	SMPL
HighStd - 1	HighStd	45648	10.0459	-7471B	PPB		31 Jul 2018 16:43:32	U	SMPL
ChkStd - 1	ChkStd	32026	7.0486	-7471B	PPB		31 Jul 2018 16:45:48	U	SMPL
PB111615BL - 1	PBS	-310	-0.0666	-7471B	PPB		31 Jul 2018 16:48:07	U	SMPL
PB111615BS - 1	LCSS	17809	3.9203	-7471B	PPB		31 Jul 2018 16:50:25	U	SMPL
J3823-01 - 1	NB-08-073018-A	4096	0.9029	-7471B	PPB		31 Jul 2018 16:52:42	U	SMPL
J3823-03 - 1	NB-08-073018-B	6592	1.4521	-7471B	PPB		31 Jul 2018 16:55:00	U	SMPL
J4231-01 - 1	2018-NYSEG-CLARK-AREA-14-1	5396	1.1889	-7471B	PPB		31 Jul 2018 16:57:18	U	SMPL
J4231-03 - 1	2018-NYSEG-CLARK-AREA-14-2	2205	0.4868	-7471B	PPB		31 Jul 2018 16:59:35	U	SMPL
J4231-05 - 1	2018-NYSEG-CLARK-AREA-14-3	13744	3.0258	-7471B	PPB		31 Jul 2018 17:01:51	U	SMPL
CCV59 - 1	CCV59	21885	4.8171	-7471B	PPB		31 Jul 2018 17:04:07	U	SMPL
CCB59 - 1	CCB59	-193	-0.0409	-7471B	PPB		31 Jul 2018 17:06:23	U	SMPL
J4231-07 - 1	2018-NYSEG-CLARK-AREA-14-4	49751	10.9488	-7471B	PPB		31 Jul 2018 17:08:40	U	SMPL
J4231-09 - 1	2018-NYSEG-CLARK-AREA-14-5	37476	8.2478	-7471B	PPB		31 Jul 2018 17:10:56	U	SMPL
J4231-11 - 1	2018-NYSEG-CLARK-AREA-14-6	91953	20.2349	-7471B	PPB		31 Jul 2018 17:13:14	U	SMPL
J4231-13 - 1	2018-NYSEG-CLARK-AREA-14-7	8881	1.9557	-7471B	PPB		31 Jul 2018 17:15:32	U	SMPL
J4231-15 - 1	2018-NYSEG-CLARK-AREA-14-8	3171	0.6993	-7471B	PPB		31 Jul 2018 17:17:52	U	SMPL
J4231-17 - 1	2018-NYSEG-CLARK-AREA-14-9	8012	1.7645	-7471B	PPB		31 Jul 2018 17:20:09	U	SMPL
J4231-19 - 1	2018-NYSEG-CLARK-AREA-14-10	9844	2.1676	-7471B	PPB		31 Jul 2018 17:22:26	U	SMPL
J4231-21 - 1	2018-NYSEG-CLARK-AREA-7	3953	0.8714	-7471B	PPB		31 Jul 2018 17:24:44	U	SMPL
J4234-01 - 1	OILY-STONE-CONCRETE	354	0.0795	-7471B	PPB		31 Jul 2018 17:27:02	U	SMPL
J4236-01 - 1	FK-GF-01-010-A	2531	0.5585	-7471B	PPB		31 Jul 2018 17:29:18	U	SMPL
CCV60 - 1	CCV60	21995	4.8413	-7471B	PPB		31 Jul 2018 17:31:35	U	SMPL
CCB60 - 1	CCB60	-161	-0.0339	-7471B	PPB		31 Jul 2018 17:33:50	U	SMPL
J4236-03 - 1	FK-GF-01-010-B	2634	0.5812	-7471B	PPB		31 Jul 2018 17:36:08	U	SMPL
J4237-01 - 1	FK-GF-02-023-A	1985	0.4383	-7471B	PPB		31 Jul 2018 17:38:24	U	SMPL
J4237-03 - 1	FK-GF-02-023-B	2148	0.4742	-7471B	PPB		31 Jul 2018 17:45:59	U	SMPL
J4237-03DUP - 1	FK-GF-02-023-BDUP	2168	0.4786	-7471B	PPB		31 Jul 2018 17:48:15	U	SMPL
J4237-03MS - 1	FK-GF-02-023-BMS	19435	4.2780	-7471B	PPB		31 Jul 2018 17:50:32	U	SMPL
J4237-03MSD - 1	FK-GF-02-023-BMSD	18895	4.1592	-7471B	PPB		31 Jul 2018 17:52:49	U	SMPL
PB111616BL - 1	PBS	-179	-0.0378	-7471B	PPB		31 Jul 2018 17:55:06	U	SMPL
PB111616BS - 1	LCSS	15829	3.4846	-7471B	PPB		31 Jul 2018 17:57:24	U	SMPL
J3825-11 - 1	CL-01-073018-A	1351	0.2988	-7471B	PPB		31 Jul 2018 17:59:41	U	SMPL
J3825-11DUP - 1	CL-01-073018-ADUP	1207	0.2672	-7471B	PPB		31 Jul 2018 18:01:58	U	SMPL
CCV61 - 1	CCV61	22138	4.8728	-7471B	PPB		31 Jul 2018 18:04:14	U	SMPL
CCB61 - 1	CCB61	-144	-0.0301	-7471B	PPB		31 Jul 2018 18:06:30	U	SMPL
J3825-11MS - 1	CL-01-073018-AMS	17837	3.9264	-7471B	PPB		31 Jul 2018 18:08:48	U	SMPL
J3825-11MSD - 1	CL-01-073018-AMSD	18059	3.9753	-7471B	PPB		31 Jul 2018 18:11:04	U	SMPL
J3825-13 - 1	CL-01-073018-B	947	0.2099	-7471B	PPB		31 Jul 2018 18:13:21	U	SMPL
J3825-15 - 1	CL-01-073018-C	988	0.2190	-7471B	PPB		31 Jul 2018 18:15:38	U	SMPL
J3825-17 - 1	CL-01-073018-D	1060	0.2348	-7471B	PPB		31 Jul 2018 18:17:54	U	SMPL
J3825-19 - 1	CL-01-073018-E	1673	0.3697	-7471B	PPB		31 Jul 2018 18:20:11	U	SMPL
J4237-03LX5 - 1		416	0.0931	-7471B	PPB		31 Jul 2018 18:22:29	U	SMPL
J4237-03A - 1		20265	4.4607	-7471B	PPB		31 Jul 2018 18:24:46	U	SMPL
J3825-11LX5 - 1		107	0.0251	-7471B	PPB		31 Jul 2018 18:27:04	U	SMPL
J3825-11A - 1		19418	4.2743	-7471B	PPB		31 Jul 2018 18:29:20	U	SMPL
CCV62 - 1	CCV62	22182	4.8825	-7471B	PPB		31 Jul 2018 18:31:36	U	SMPL
CCB62 - 1	CCB62	-161	-0.0339	-7471B	PPB		31 Jul 2018 18:33:53	U	SMPL
J4231-07DLX2 - 1		23764	5.2306	-7471B	PPB		31 Jul 2018 18:36:10	U	SMPL
J4231-11DLX5 - 1		17225	3.7918	-7471B	PPB		31 Jul 2018 18:38:26	U	SMPL
CCV63 - 1	CCV63	21808	4.8002	-7471B	PPB		31 Jul 2018 18:40:43	U	SMPL
CCB63 - 1	CCB63	-195	-0.0413	-7471B	PPB		31 Jul 2018 18:43:00	U	SMPL