



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97021

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	07/31/2018	15:10
2	CAL2	1	Water			22.0	7.00	07/31/2018	15:12
3	CAL3	1	Water			21.9	10.04	07/31/2018	15:14
4	ICV	1	Water			22.1	7.00	07/31/2018	15:16
5	CCV1	1	Water			21.9	2.00	07/31/2018	15:18
6	J4250-02	1	Water			27.0	7.17	07/31/2018	15:20
7	J4250-02DUP	1	Water			27.1	7.18	07/31/2018	15:22
8	J4252-02	1	Water			26.4	7.35	07/31/2018	15:24
9	CCV2	1	Water			22.0	12.00	07/31/2018	15:26

LB 97021

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity073118-1

WorkList ID : 115122

Date : 7/31/2018 2:51:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/31/2018	Water	J4250-02	Corrosivity	Cool 4 deg C	PSEG01	K63	2576	07/31/2018	9040C
07/31/2018	Water	J4252-02	Corrosivity	Cool 4 deg C	PSEG01	K63	FM-1026	07/31/2018	9040C

Date/Time 07/31/18 3:00

Received by: SP

Relinquished by: CP

Date/Time 07/31/18 3:30

Received by: CP

Relinquished by: CP