



Analytical Summary Report

Analysis Method: 9045C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97022

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	07/31/2018	15:10
2	CAL2	1	Water			22.0	7.00	07/31/2018	15:12
3	CAL3	1	Water			21.9	10.04	07/31/2018	15:14
4	ICV	1	Water			22.1	7.00	07/31/2018	15:16
5	CCV1	1	Water			21.8	2.00	07/31/2018	16:15
6	J3826-06	1	Solid	20.02	20	26.8	7.61	07/31/2018	16:17
7	J3826-06DUP	1	Solid	20.03	20	26.9	7.62	07/31/2018	16:19
8	J4241-04	1	Solid	20.02	20	25.9	8.42	07/31/2018	16:21
9	J4251-01	1	Solid	20.02	20	25.2	5.73	07/31/2018	16:23
10	CCV2	1	Water			22.1	12.00	07/31/2018	16:25

WORKLIST(Hardcopy Internal Chain)

LB 97022

WorkList Name : corrosivity073118

WorkList ID : 115121

Date : 7/31/2018 2:50:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/31/2018	Solid	J3826-06	Corrosivity	Cool 4 deg C	PSEG05		OK-03-073118	07/31/2018	9045C
07/31/2018	Solid	J4241-04	Corrosivity	Cool 4 deg C	PSEG01	K63	FS-RO-226	07/31/2018	9045C
07/30/2018	Solid	J4251-01	Corrosivity	Cool 4 deg C	USAR03	L23	CB-TCLP-11.4-0718	07/30/2018	9045C

Date/Time 07/31/18 4:05

Received by: SP

Relinquished by: SP

Date/Time 07/31/18 4:35

Received by: CP

Relinquished by: SP