



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97024

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	08/01/2018	09:50
2	J4250-02		1	8	>100.0	>212.0	>212.0	08/01/2018	10:20
3	J4250-02DUP		1	8	>100.0	>212.0	>212.0	08/01/2018	10:45
4	J4252-02		1	7	>100.0	>212.0	>212.0	08/01/2018	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB97024

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp8118 WorkList ID : 115142 Date : 8/1/2018 9:00:08 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/03/2018	Water	J4250-02 C	Flash Point	Cool 4 deg C	PSEG01	K63	2576	07/31/2018	1010A
08/03/2018	Water	J4252-02 C	Flash Point	Cool 4 deg C	PSEG01	K63	FM-1026	07/31/2018	1010A

Date/Time 08/01/2018 09:10 AM
Received by: KS
Relinquished by: CP

Date/Time 08/01/2018 11:25 AM
Received by: CP
Relinquished by: KS