

LB97027


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4944E



A= 0.0000e+000

B= 2.0300e-004

C= -4.9695e-003

Rho= 0.9998660

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	5	0.000	5				
0.2	0.200	0.204	0.004	1027	0.0 %	1027				
2.5	2.500	2.472	-0.028	12204	0.0 %	12204				
5.0	5.000	5.098	0.098	25137	0.0 %	25137				
7.5	7.500	7.398	-0.102	36466	0.0 %	36466				
10.0	10.000	10.032	0.032	49445	0.0 %	49445				

LB97027 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Aug 2018 11:26:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	5	-	0.00007471B	PPB		01 Aug 2018 11:34:41	S	Std
0.2 - 1	50.2	1027	-	0.20007471B	PPB		01 Aug 2018 11:36:59	S	Std
2.5 - 1	52.5	12204	-	2.50007471B	PPB		01 Aug 2018 11:39:16	S	Std
5.0 - 1	55	25137	-	5.00007471B	PPB		01 Aug 2018 11:41:34	S	Std
7.5 - 1	57.5	36466	-	7.50007471B	PPB		01 Aug 2018 11:46:23	S	Std
10.0 - 1	510	49445	-	10.00007471B	PPB		01 Aug 2018 11:51:13	S	Std
ICV83 - 1	ICV83	18304	3.7108	-7471B	PPB		01 Aug 2018 11:53:52	U	SMPL
ICB83 - 1	ICB83	-178	-0.0411	-7471B	PPB		01 Aug 2018 11:56:08	U	SMPL
CCV68 - 1	CCV68	25409	5.1531	-7471B	PPB		01 Aug 2018 11:58:25	U	SMPL
CCB68 - 1	CCB68	-77	-0.0206	-7471B	PPB		01 Aug 2018 12:00:41	U	SMPL
CRA - 1	CRA	1005	0.1990	-7471B	PPB		01 Aug 2018 12:02:58	U	SMPL
HighStd - 1	HighStd	49889	10.1226	-7471B	PPB		01 Aug 2018 12:05:14	U	SMPL
ChkStd - 1	ChkStd	29748	6.0339	-7471B	PPB		01 Aug 2018 12:07:30	U	SMPL
PB111654BL - 1	PBS	196	0.0348	-7471B	PPB		01 Aug 2018 12:13:24	U	SMPL
PB111654BS - 1	LCSS	17422	3.5317	-7471B	PPB		01 Aug 2018 12:15:40	U	SMPL
J3826-05 - 1	OK-03-073118	4746	0.9585	-7471B	PPB		01 Aug 2018 12:17:57	U	SMPL
J3826-05DUP - 1	OK-03-073118DUP	4186	0.8448	-7471B	PPB		01 Aug 2018 12:22:47	U	SMPL
J3826-05MS - 1	OK-03-073118MS	21576	4.3750	-7471B	PPB		01 Aug 2018 12:25:04	U	SMPL
J3826-05MSD - 1	OK-03-073118MSD	21737	4.4077	-7471B	PPB		01 Aug 2018 12:27:21	U	SMPL
J4085-01 - 1	1522	815892	HIGH	-7471B	PPB		01 Aug 2018 12:29:39	X	
CCV69 - 1	CCV69	23897	4.8462	-7471B	PPB		01 Aug 2018 12:31:28	U	SMPL
CCB69 - 1	CCB69	-151	-0.0356	-7471B	PPB		01 Aug 2018 12:34:32	U	SMPL
J4238-01 - 1	RVB-1	746	0.1465	-7471B	PPB		01 Aug 2018 12:36:51	U	SMPL
J4238-02 - 1	RVB-3	975	0.1930	-7471B	PPB		01 Aug 2018 12:39:06	U	SMPL
J4238-03 - 1	RVB-9	1518	0.3032	-7471B	PPB		01 Aug 2018 12:41:22	U	SMPL
J4241-01 - 1	FS-RO-226	17517	3.5510	-7471B	PPB		01 Aug 2018 12:43:38	U	SMPL
J4257-01 - 1	FK-BPM-03-001-A	2974	0.5988	-7471B	PPB		01 Aug 2018 12:45:55	U	SMPL
J4257-03 - 1	FK-BPM-03-001-B	2861	0.5758	-7471B	PPB		01 Aug 2018 12:48:13	U	SMPL
J4257-05 - 1	FK-BPM-03-001-C	1984	0.3978	-7471B	PPB		01 Aug 2018 12:50:31	U	SMPL
J4257-07 - 1	FK-BPM-03-001-D	1952	0.3913	-7471B	PPB		01 Aug 2018 12:52:48	U	SMPL
J3826-05LX5 - 1		1001	0.1982	-7471B	PPB		01 Aug 2018 12:55:06	U	SMPL
CCV70 - 1	CCV70	25000	5.0701	-7471B	PPB		01 Aug 2018 12:59:39	U	SMPL
CCB70 - 1	CCB70	18	-0.0013	-7471B	PPB		01 Aug 2018 13:01:55	U	SMPL
J4085-01DLX10 - 1	1522	207121	42.0409	-7471B	PPB		01 Aug 2018 13:14:22	U	SMPL
J3826-05A - 1		20833	4.2242	-7471B	PPB		01 Aug 2018 13:19:34	U	SMPL
J4085-01DL2X100 - 1		21529	4.3654	-7471B	PPB		01 Aug 2018 13:24:29	U	SMPL
CCV71 - 1	CCV71	24310	4.9300	-7471B	PPB		01 Aug 2018 13:29:05	U	SMPL
CCB71 - 1	CCB71	-64	-0.0180	-7471B	PPB		01 Aug 2018 13:31:23	U	SMPL