



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB97032

Analyst By : sweta
Supervisor Review By : apatel
Slope : 98.8
pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	08/01/2018	13:40
2	CAL2	1	Water			22.4	7.00	08/01/2018	13:42
3	CAL3	1	Water			22.8	10.04	08/01/2018	13:44
4	ICV	1	Water			22.5	7.00	08/01/2018	13:46
5	CCV1	1	Water			22.6	2.00	08/01/2018	13:48
6	J4279-04	1	Water			25.1	9.32	08/01/2018	13:50
7	J4279-04DUP	1	Water			25.2	9.33	08/01/2018	13:52
8	J4279-05	1	Water			25.3	7.81	08/01/2018	13:54
9	CCV2	1	Water			22.6	12.00	08/01/2018	13:56

LB 97032

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity080118-1

WorkList ID : 115162

Date : 8/1/2018 1:22:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/14/2018	Water	J4279-04 C	Corrosivity	Cool 4 deg C	TETR06	O52	WE13-GW-TANK1-IDW-07291	08/29/2018	9040C
08/14/2018	Water	J4279-05 C	Corrosivity	Cool 4 deg C	TETR06	O52	WE13-GW-TANK2-IDW-07291	08/29/2018	9040C

Date/Time 08/01/18 1:30
Received by: SP
Relinquished by: CP

Date/Time 08/01/18 2:10
Received by: CP
Relinquished by: SP