


7470A

LB 97035
INSTRUMENT ID CV1

Linear



A= 0.0000e+000

B= 2.6473e-004

C= -5.0182e-003

Rho= 0.9997921

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	-7	0.000	-7				
0.2	0.200	0.179	-0.021	694	0.0 %	694				
2.5	2.500	2.492	-0.008	9433	0.0 %	9433				
5.0	5.000	5.001	0.001	18911	0.0 %	18911				
7.5	7.500	7.645	0.145	28899	0.0 %	28899				
10.0	10.000	9.889	-0.111	37375	0.0 %	37375				

LB97035 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Aug 2018 15:18:34

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-7	-	0.00007470A	PPB		01 Aug 2018 15:28:11	S	Std
0.2 - 1	SO 2	694	-	0.20007470A	PPB		01 Aug 2018 15:30:19	S	Std
2.5 - 1	SO 5	9433	-	2.50007470A	PPB		01 Aug 2018 15:32:27	S	Std
5.0 - 1	SO 5	18911	-	5.00007470A	PPB		01 Aug 2018 15:34:35	S	Std
7.5 - 1	SO 5	28899	-	7.50007470A	PPB		01 Aug 2018 15:36:43	S	Std
10.0 - 1	SO 5	37375	-	10.00007470A	PPB		01 Aug 2018 15:38:51	S	Std
ICV84 - 1	ICV84	14591	3.8577	-7470A	PPB		01 Aug 2018 15:41:28	U	SMPL
ICB84 - 1	ICB84	22	0.0008	-7470A	PPB		01 Aug 2018 15:46:07	U	SMPL
CCV72 - 1	CCV72	19273	5.0971	-7470A	PPB		01 Aug 2018 15:48:14	U	SMPL
CCB72 - 1	CCB72	-98	-0.0310	-7470A	PPB		01 Aug 2018 15:50:20	U	SMPL
CRA - 1	CRA	774	0.1999	-7470A	PPB		01 Aug 2018 15:59:29	U	SMPL
HighStd - 1	HighStd	36947	9.7760	-7470A	PPB		01 Aug 2018 16:01:36	U	SMPL
ChkStd - 1	ChkStd	26643	7.0482	-7470A	PPB		01 Aug 2018 16:03:43	U	SMPL
PB111671BL - 1	PBW	-101	-0.0318	-7470A	PPB		01 Aug 2018 16:08:16	U	SMPL
PB111671BS - 1	LCSW	14376	3.8007	-7470A	PPB		01 Aug 2018 16:10:22	U	SMPL
PB111671TB - 1	PB111648TB	362	0.0908	-7470A	PPB		01 Aug 2018 16:19:13	U	SMPL
J3826-06 - 1	OK-03-073118	363	0.0911	-7470A	PPB		01 Aug 2018 16:21:20	U	SMPL
J4241-04 - 1	FS-RO-226	235	0.0572	-7470A	PPB		01 Aug 2018 16:23:28	U	SMPL
J4241-04DUP - 1	FS-RO-226DUP	251	0.0614	-7470A	PPB		01 Aug 2018 16:25:37	U	SMPL
J4241-04MS - 1	FS-RO-226MS	14546	3.8457	-7470A	PPB		01 Aug 2018 16:27:43	U	SMPL
CCV73 - 1	CCV73	19249	5.0908	-7470A	PPB		01 Aug 2018 16:29:50	U	SMPL
CCB73 - 1	CCB73	242	0.0590	-7470A	PPB		01 Aug 2018 16:31:57	U	SMPL
J4241-04MSD - 1	FS-RO-226MSD	14772	3.9056	-7470A	PPB		01 Aug 2018 16:36:31	U	SMPL
J4251-01 - 1	CB-TCLP-11-4-0718	325	0.0810	-7470A	PPB		01 Aug 2018 16:38:38	U	SMPL
PB111672BL - 1	PBW	96	0.0204	-7470A	PPB		01 Aug 2018 16:40:45	U	SMPL
PB111672BS - 1	LCSW	14510	3.8362	-7470A	PPB		01 Aug 2018 16:42:52	U	SMPL
PB111672TB - 1	PB111652TB	238	0.0580	-7470A	PPB		01 Aug 2018 16:44:59	U	SMPL
J4250-02 - 1	2576	229	0.0556	-7470A	PPB		01 Aug 2018 16:47:07	U	SMPL
J4250-02DUP - 1	2576DUP	231	0.0561	-7470A	PPB		01 Aug 2018 16:49:15	U	SMPL
J4250-02MS - 1	2576MS	14729	3.8942	-7470A	PPB		01 Aug 2018 16:51:23	U	SMPL
J4250-02MSD - 1	2576MSD	15078	3.9866	-7470A	PPB		01 Aug 2018 16:53:32	U	SMPL
J4252-02 - 1	FM-1026	187	0.0445	-7470A	PPB		01 Aug 2018 16:58:17	U	SMPL
CCV74 - 1	CCV74	19424	5.1371	-7470A	PPB		01 Aug 2018 17:00:23	U	SMPL
CCB74 - 1	CCB74	184	0.0437	-7470A	PPB		01 Aug 2018 17:02:30	U	SMPL
PB111673BL - 1	PBW	175	0.0413	-7470A	PPB		01 Aug 2018 17:07:44	U	SMPL
PB111673BS - 1	LCSW	13154	3.4772	-7470A	PPB		01 Aug 2018 17:09:51	U	SMPL
J4085-05 - 1	1511B	25508	6.7477	-7470A	PPB		01 Aug 2018 17:11:57	U	SMPL
J4250-01 - 1	2576	-105	-0.0328	-7470A	PPB		01 Aug 2018 17:14:05	U	SMPL
J4250-01DUP - 1	2576DUP	147	0.0339	-7470A	PPB		01 Aug 2018 17:16:15	U	SMPL
J4250-01MS - 1	2576MS	15191	4.0165	-7470A	PPB		01 Aug 2018 17:18:22	U	SMPL
J4250-01MSD - 1	2576MSD	14857	3.9281	-7470A	PPB		01 Aug 2018 17:20:30	U	SMPL
J4252-01 - 1	FM-1026	-55	-0.0196	-7470A	PPB		01 Aug 2018 17:22:38	U	SMPL
CCV75 - 1	CCV75	19353	5.1183	-7470A	PPB		01 Aug 2018 17:29:03	U	SMPL
CCB75 - 1	CCB75	95	0.0201	-7470A	PPB		01 Aug 2018 17:31:13	U	SMPL
J4241-04LX5 - 1		207	0.0498	-7470A	PPB		01 Aug 2018 17:33:21	U	SMPL
J4241-04A - 1		14687	3.8831	-7470A	PPB		01 Aug 2018 17:37:56	U	SMPL
J4250-02LX5 - 1		69	0.0132	-7470A	PPB		01 Aug 2018 17:40:03	U	SMPL
J4250-02A - 1		14697	3.8857	-7470A	PPB		01 Aug 2018 17:42:12	U	SMPL
J4250-01LX5 - 1		3	-0.0042	-7470A	PPB		01 Aug 2018 17:44:19	U	SMPL
J4250-01A - 1		15063	3.9826	-7470A	PPB		01 Aug 2018 17:46:27	U	SMPL
CCV76 - 1	CCV76	18908	5.0005	-7470A	PPB		01 Aug 2018 17:48:35	U	SMPL
CCB76 - 1	CCB76	34	0.0040	-7470A	PPB		01 Aug 2018 17:50:44	U	SMPL