



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97042

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/01/2018	15:40
2	CAL2	1	Water			21.8	7.01	08/01/2018	15:42
3	CAL3	1	Water			21.6	10.04	08/01/2018	15:44
4	ICV	1	Water			21.7	7.00	08/01/2018	15:46
5	CCV1	1	Water			21.8	2.00	08/01/2018	15:48
6	J4284-03	1	Water			27.1	7.36	08/01/2018	15:50
7	J4284-03DUP	1	Water			27.2	7.37	08/01/2018	15:52
8	J4287-01	1	Water			27.2	6.69	08/01/2018	15:54
9	CCV2	1	Water			21.8	12.00	08/01/2018	15:56

WORKLIST(Hardcopy Internal Chain)

LB 97042

WorkList Name : corrosivity080118-2

WorkList ID : 115165

Date : 8/1/2018 3:21:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/01/2018	Water	J4284-03	Corrosivity	Cool 4 deg C	PSEG01	K63	OILY-WATER	08/01/2018	9040C
08/01/2018	Water	J4287-01	Corrosivity	Cool 4 deg C	ATCE02	O63	A4742	08/01/2018	9040C

3:30

Date/Time 08/01/18 8:30

Received by: Sp

Relinquished by: CP

Date/Time 08/01/18 5:00

Received by: CP

Relinquished by: CP Sp