



Analytical Summary Report

Analysis Method: 9045C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97043

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/01/2018	15:40
2	CAL2	1	Water			21.8	7.01	08/01/2018	15:42
3	CAL3	1	Water			21.6	10.04	08/01/2018	15:44
4	ICV	1	Water			21.7	7.00	08/01/2018	15:46
5	CCV1	1	Water			21.9	2.00	08/01/2018	16:20
6	J4260-03	1	Solid	20.02	20	26.9	9.12	08/01/2018	16:22
7	J4260-03DUP	1	Solid	20.04	20	27.0	9.13	08/01/2018	15:24
8	J4278-02	1	Solid	20.02	20	26.9	8.41	08/01/2018	16:26
9	J4278-04	1	Solid	20.03	20	26.8	8.66	08/01/2018	16:28
10	J4278-06	1	Solid	20.02	20	26.9	8.25	08/01/2018	16:30
11	J4285-04	1	Solid	20.04	20	27.1	8.51	08/01/2018	16:32
12	J4285-08	1	Solid	20.02	20	27.1	7.76	08/01/2018	16:34
13	J4285-12	1	Solid	20.03	20	27.1	9.52	08/01/2018	16:36
14	J4285-16	1	Solid	20.02	20	27.1	8.15	08/01/2018	16:38
15	J4285-20	1	Solid	20.04	20	27.1	6.98	08/01/2018	16:40
16	CCV2	1	Water			22.0	12.00	08/01/2018	16:42
17	J4285-24	1	Solid	20.02	20	27.2	9.79	08/01/2018	16:44
18	CCV3	1	Water			22.1	2.00	08/01/2018	16:46

WORKLIST(Hardcopy Internal Chain)

LB 94043

WorkList Name : corrosivity080118

WorkList ID : 115164

Date : 8/1/2018 3:20:50 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/01/2018	Solid	J4260-03	Corrosivity	Cool 4 deg C	PSEG01	K63	TS-LINE-RO-135	08/01/2018	9045C
08/01/2018	Solid	J4278-02	Corrosivity	Cool 4 deg C	PSEG05	K63	EO-03-080118-A	08/01/2018	9045C
08/01/2018	Solid	J4278-04	Corrosivity	Cool 4 deg C	PSEG05	K63	EO-03-080118-B	08/01/2018	9045C
08/01/2018	Solid	J4278-06	Corrosivity	Cool 4 deg C	PSEG05	K63	EO-03-080118-C	08/01/2018	9045C
08/01/2018	Solid	J4285-04	Corrosivity	Cool 4 deg C	PSEG01	K63	KG-PIT-1	08/01/2018	9045C
08/01/2018	Solid	J4285-08	Corrosivity	Cool 4 deg C	PSEG01	K63	KG-PIT-2	08/01/2018	9045C
08/01/2018	Solid	J4285-12	Corrosivity	Cool 4 deg C	PSEG01	K63	KG-PIT-3	08/01/2018	9045C
08/01/2018	Solid	J4285-16	Corrosivity	Cool 4 deg C	PSEG01	K63	KG-PIT-4	08/01/2018	9045C
08/01/2018	Solid	J4285-20	Corrosivity	Cool 4 deg C	PSEG01	K63	KG-PIT-5	08/01/2018	9045C
08/01/2018	Solid	J4285-24	Corrosivity	Cool 4 deg C	PSEG01	K63	KG-PIT-6	08/01/2018	9045C

Date/Time 08/01/18 4:00

Received by: SP

Relinquished by: CP

Date/Time 08/01/18 5:00

Received by: CP

Relinquished by: SP