


LB97047

245.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.8889e-004

C= 2.5419e-003

Rho= 0.9999135

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	13	0.000	13				
0.05	0.050	0.060	0.010	199	0.0 %	199				
0.2	0.200	0.200	-0.000	682	0.0 %	682				
2.5	2.500	2.527	0.027	8739	0.0 %	8739				
5.0	5.000	4.980	-0.020	17230	0.0 %	17230				
7.5	7.500	7.401	-0.099	25611	0.0 %	25611				
10.0	10.000	10.075	0.075	34867	0.0 %	34867				

LB97047 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 Aug 2018 18:54:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	13	-	0.0000	245.1	PPB	01 Aug 2018 18:55:22	S	Std
0.05 - 1	50.5	199	-	0.0500	245.1	PPB	01 Aug 2018 18:57:31	S	Std
0.2 - 1	50.2	682	-	0.2000	245.1	PPB	01 Aug 2018 18:59:39	S	Std
2.5 - 1	52.5	8739	-	2.5000	245.1	PPB	01 Aug 2018 19:01:47	S	Std
5.0 - 1	55	17230	-	5.0000	245.1	PPB	01 Aug 2018 19:03:55	S	Std
7.5 - 1	57.5	25611	-	7.5000	245.1	PPB	01 Aug 2018 19:09:06	S	Std
10.0 - 1	510	34867	-	10.0000	245.1	PPB	01 Aug 2018 19:13:38	S	Std
ICV85 - 1	ICV85	14092	4.0736	-245.1		PPB	01 Aug 2018 19:18:35	U	SMPL
ICB85 - 1	ICB85	-81	-0.0209	-245.1		PPB	01 Aug 2018 19:20:41	U	SMPL
CCV77 - 1	CCV77	17217	4.9764	-245.1		PPB	01 Aug 2018 19:22:48	U	SMPL
CCB77 - 1	CCB77	-132	-0.0356	-245.1		PPB	01 Aug 2018 19:24:55	U	SMPL
CRA - 1	CRA	187	0.0566	-245.1		PPB	01 Aug 2018 19:27:03	U	SMPL
HighStd - 1	HighStd	34971	10.1054	-245.1		PPB	01 Aug 2018 19:29:09	U	SMPL
ChkStd - 1	ChkStd	24538	7.0914	-245.1		PPB	01 Aug 2018 19:31:16	U	SMPL
PB111676BL - 1	PBW	28	0.0106	-245.1		PPB	01 Aug 2018 19:35:57	U	SMPL
PB111676BS - 1	LCSW	12948	3.7431	-245.1		PPB	01 Aug 2018 19:38:04	U	SMPL
J4084-04 - 1	1511B	23794	6.8765	-245.1		PPB	01 Aug 2018 19:46:41	U	SMPL
CCV78 - 1	CCV78	17321	5.0065	-245.1		PPB	01 Aug 2018 19:48:49	U	SMPL
CCB78 - 1	CCB78	-178	-0.0489	-245.1		PPB	01 Aug 2018 19:50:58	U	SMPL