



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97057

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.6	08/02/2018	09:20
2	J4284-03		1	7	>100.0	>212.0	>212.0	08/02/2018	09:50
3	J4284-03DUP		1	7	>100.0	>212.0	>212.0	08/02/2018	10:15
4	J4287-01		1	5	>100.0	>212.0	>212.0	08/02/2018	10:42

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB97057

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp8218

WorkList ID : 115200

Date : 8/2/2018 8:39:22 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/06/2018	Water	J4284-03	Flash Point	Cool 4 deg C	PSEG01	K63	OILY-WATER	08/01/2018	1010A
08/03/2018	Water	J4287-01	Flash Point	Cool 4 deg C	ATCE02	O63	A4742	08/01/2018	1010A

Date/Time 08/02/18 08:45AM

Received by: KS

Relinquished by: CP

Date/Time 08/02/18 10:45AM

Received by: CP

Relinquished by: KS