



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB97064

Reagent/Standard	Lot/Log #
conductivity standard	WP62773

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB97064BL		1	22.5	0.92	0.97	08/02/2018	13:05
2	LB97064BS	1413	1	22.4	1330.00	1399.50	08/02/2018	13:18
3	J4080-13		1	24.1	895.00	910.65	08/02/2018	13:28
4	J4080-13DUP		1	24.1	894.00	909.64	08/02/2018	13:36

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$