

Analysis Method: 120.1

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB97065

Reagent/Standard	Lot/Log #
conductivity standard	WP62773

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB97065BL		1	22.5	0.92	0.97	08/02/2018	13:05
2	LB97065BS	1413	1	22.4	1330.00	1399.50	08/02/2018	13:18
3	J4080-14		1	23.9	892.00	911.14	08/02/2018	13:42
4	J4080-14DUP		1	23.9	891.00	910.12	08/02/2018	13:50

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$