


7471B

LB97071
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44452



A= 0.0000e+000

B= 2.2589e-004

C= 4.0077e-002

Rho= 0.9998594

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.046	0.046	26	0.000	26				
0.2	0.200	0.210	0.010	752	0.0 %	752				
2.5	2.500	2.516	0.016	10962	0.0 %	10962				
5.0	5.000	4.891	-0.109	21475	0.0 %	21475				
7.5	7.500	7.455	-0.045	32826	0.0 %	32826				
10.0	10.000	10.081	0.081	44452	0.0 %	44452				

LB97071 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Aug 2018 15:07:15

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7471B	PPB	02 Aug 2018 15:15:09	S	Std
0.2 - 1	50.2	752	-	0.2000	7471B	PPB	02 Aug 2018 15:17:27	S	Std
2.5 - 1	52.5	10962	-	2.5000	7471B	PPB	02 Aug 2018 15:19:45	S	Std
5.0 - 1	55	21475	-	5.0000	7471B	PPB	02 Aug 2018 15:22:03	S	Std
7.5 - 1	57.5	32826	-	7.5000	7471B	PPB	02 Aug 2018 15:24:20	S	Std
10.0 - 1	510	44452	-	10.0000	7471B	PPB	02 Aug 2018 15:29:15	S	Std
ICV86 - 1	ICV86	17347	3.9586	-7471B	PPB	PPB	02 Aug 2018 15:31:58	U	SMPL
ICB86 - 1	ICB86	-101	0.0173	-7471B	PPB	PPB	02 Aug 2018 15:34:15	U	SMPL
CCV79 - 1	CCV79	22166	5.0472	-7471B	PPB	PPB	02 Aug 2018 15:36:32	U	SMPL
CCB79 - 1	CCB79	-119	0.0132	-7471B	PPB	PPB	02 Aug 2018 15:38:48	U	SMPL
CRA - 1	CRA	786	0.2176	-7471B	PPB	PPB	02 Aug 2018 15:41:05	U	SMPL
HighStd - 1	HighStd	45082	10.2238	-7471B	PPB	PPB	02 Aug 2018 15:43:20	U	SMPL
ChkStd - 1	ChkStd	31296	7.1096	-7471B	PPB	PPB	02 Aug 2018 15:45:36	U	SMPL
PB111685BL - 1	PBS	-303	-0.0284	-7471B	PPB	PPB	02 Aug 2018 15:47:54	U	SMPL
PB111685BS - 1	LCSS	15739	3.5954	-7471B	PPB	PPB	02 Aug 2018 15:50:12	U	SMPL
J4260-01 - 1	TS-LINER-RO-135	11210	2.5723	-7471B	PPB	PPB	02 Aug 2018 15:52:29	U	SMPL
J4278-01 - 1	EO-03-080118-A	8084	1.8662	-7471B	PPB	PPB	02 Aug 2018 15:54:46	U	SMPL
J4278-01DUP - 1	EO-03-080118-ADUP	8161	1.8836	-7471B	PPB	PPB	02 Aug 2018 15:57:05	U	SMPL
J4278-01MS - 1	EO-03-080118-AMS	25396	5.7768	-7471B	PPB	PPB	02 Aug 2018 15:59:23	U	SMPL
J4278-01MSD - 1	EO-03-080118-AMSD	24468	5.5672	-7471B	PPB	PPB	02 Aug 2018 16:01:39	U	SMPL
CCV80 - 1	CCV80	21737	4.9503	-7471B	PPB	PPB	02 Aug 2018 16:03:56	U	SMPL
CCB80 - 1	CCB80	-165	0.0028	-7471B	PPB	PPB	02 Aug 2018 16:06:12	U	SMPL
J4278-03 - 1	EO-03-080118-B	9223	2.1235	-7471B	PPB	PPB	02 Aug 2018 16:08:28	U	SMPL
J4278-05 - 1	EO-03-080118-C	4068	0.9590	-7471B	PPB	PPB	02 Aug 2018 16:10:45	U	SMPL
J4285-01 - 1	KG-PIT-1	39806	9.0320	-7471B	PPB	PPB	02 Aug 2018 16:13:01	U	SMPL
J4285-05 - 1	KG-PIT-2	1482	0.3748	-7471B	PPB	PPB	02 Aug 2018 16:15:17	U	SMPL
J4285-09 - 1	KG-PIT-3	85577	19.3713	-7471B	PPB	PPB	02 Aug 2018 16:17:36	U	SMPL
J4285-13 - 1	KG-PIT-4	433854	HIGH	-7471B	PPB	PPB	02 Aug 2018 16:19:53	X	
J4285-17 - 1	KG-PIT-5	68703	15.5596	-7471B	PPB	PPB	02 Aug 2018 16:21:46	U	SMPL
J4285-21 - 1	KG-PIT-6	34362	7.8022	-7471B	PPB	PPB	02 Aug 2018 16:24:20	U	SMPL
J4278-01LX5 - 1		1312	0.3364	-7471B	PPB	PPB	02 Aug 2018 16:26:41	U	SMPL
J4278-01A - 1		27244	6.1943	-7471B	PPB	PPB	02 Aug 2018 16:32:00	U	SMPL
CCV81 - 1	CCV81	22207	5.0565	-7471B	PPB	PPB	02 Aug 2018 16:34:16	U	SMPL
CCB81 - 1	CCB81	-138	0.0089	-7471B	PPB	PPB	02 Aug 2018 16:36:34	U	SMPL
J4285-09DLX5 - 1		17612	4.0185	-7471B	PPB	PPB	02 Aug 2018 16:38:51	U	SMPL
J4285-13DLX10 - 1		67849	15.3667	-7471B	PPB	PPB	02 Aug 2018 16:41:07	U	SMPL
J4285-17DLX5 - 1		13563	3.1039	-7471B	PPB	PPB	02 Aug 2018 16:43:23	U	SMPL
J4285-13DL2X50 - 1		13744	3.1447	-7471B	PPB	PPB	02 Aug 2018 16:45:42	U	SMPL
CCV82 - 1	CCV82	22440	5.1091	-7471B	PPB	PPB	02 Aug 2018 16:47:59	U	SMPL
CCB82 - 1	CCB82	-204	-0.0060	-7471B	PPB	PPB	02 Aug 2018 16:50:16	U	SMPL