



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97085

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.9	4.01	08/02/2018	15:05
2	CAL2	1	Water			22.0	7.01	08/02/2018	15:07
3	CAL3	1	Water			21.7	10.04	08/02/2018	15:09
4	ICV	1	Water			21.8	7.00	08/02/2018	15:11
5	CCV1	1	Water			21.8	2.00	08/02/2018	17:00
6	J4316-09	1	Water			26.0	7.77	08/02/2018	17:02
7	J4316-09DUP	1	Water			26.1	7.78	08/02/2018	17:04
8	J4316-12	1	Water			25.2	7.75	08/02/2018	17:06
9	J4316-13	1	Water			25.1	7.86	08/02/2018	17:08
10	J4316-14	1	Water			25.1	7.81	08/02/2018	17:10
11	J4317-01	1	Water			25.8	7.62	08/02/2018	17:12
12	CCV2	1	Water			21.9	12.00	08/02/2018	17:14

WORKLIST(Hardcopy Internal Chain)

LB 97085

WorkList Name : pH080218

WorkList ID : 115221

Date : 8/2/2018 4:43:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/02/2018	Water	J4316-09	pH	Cool 4 deg C	PSEG01	K63	OW-2-4	08/02/2018	9040C
08/02/2018	Water	J4316-12	pH	Cool 4 deg C	PSEG01	K63	OW-2	08/02/2018	9040C
08/02/2018	Water	J4316-13	pH	Cool 4 deg C	PSEG01	K63	OW-3	08/02/2018	9040C
08/02/2018	Water	J4316-14	pH	Cool 4 deg C	PSEG01	K63	OW-4	08/02/2018	9040C
08/02/2018	Water	J4317-01	pH	Cool 4 deg C	PSEG01	K63	N28424	08/02/2018	9040C

Date/Time 08/02/18 4:50
 Received by: SP
 Relinquished by: CP

Date/Time 08/02/18 5:30
 Received by: CP
 Relinquished by: SP