


L397093

7470A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.7485e-004

C= 1.2121e-002

Rho= 0.9996830

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	7	0.000	7				
0.2	0.200	0.195	-0.005	664	0.0 %	664				
2.5	2.500	2.450	-0.050	8870	0.0 %	8870				
5.0	5.000	4.975	-0.025	18057	0.0 %	18057				
7.5	7.500	7.685	0.185	27917	0.0 %	27917				
10.0	10.000	9.881	-0.119	35906	0.0 %	35906				

LB97093 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Aug 2018 12:21:51

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	7	-	0.0000	7470A	PPB	03 Aug 2018 12:24:28	S	Std
0.2 - 1	50.2	664	-	0.2000	7470A	PPB	03 Aug 2018 12:26:36	S	Std
2.5 - 1	52.5	8870	-	2.5000	7470A	PPB	03 Aug 2018 12:28:44	S	Std
5.0 - 1	55	18057	-	5.0000	7470A	PPB	03 Aug 2018 12:30:53	S	Std
7.5 - 1	57.5	27917	-	7.5000	7470A	PPB	03 Aug 2018 12:33:01	S	Std
10.0 - 1	50	35906	-	10.0000	7470A	PPB	03 Aug 2018 12:35:09	S	Std
ICV88 - 1	ICV88	14455	3.9851	-	7470A	PPB	03 Aug 2018 12:37:42	U	SMPL
ICB88 - 1	ICB88	-153	-0.0299	-	7470A	PPB	03 Aug 2018 12:39:48	U	SMPL
CCV90 - 1	CCV90	18424	5.0760	-	7470A	PPB	03 Aug 2018 12:41:56	U	SMPL
CCB90 - 1	CCB90	-131	-0.0239	-	7470A	PPB	03 Aug 2018 12:44:02	U	SMPL
CRA - 1	CRA	602	0.1776	-	7470A	PPB	03 Aug 2018 12:46:10	U	SMPL
HighStd - 1	HighStd	36024	9.9134	-	7470A	PPB	03 Aug 2018 12:48:17	U	SMPL
ChkStd - 1	ChkStd	24415	6.7226	-	7470A	PPB	03 Aug 2018 12:50:24	U	SMPL
PB111707BL - 1	PBW	-258	-0.0588	-	7470A	PPB	03 Aug 2018 12:52:33	U	SMPL
PB111707BS - 1	LCSW	12857	3.5459	-	7470A	PPB	03 Aug 2018 12:54:42	U	SMPL
PB111707TB - 1	PB111678TB	-115	-0.0195	-	7470A	PPB	03 Aug 2018 12:56:49	U	SMPL
J4260-03 - 1	TS-LINER-RO-135	14	0.0160	-	7470A	PPB	03 Aug 2018 12:58:57	U	SMPL
J4278-02 - 1	EO-03-080118-A	-19	0.0069	-	7470A	PPB	03 Aug 2018 13:01:06	U	SMPL
J4278-04 - 1	EO-03-080118-B	17	0.0168	-	7470A	PPB	03 Aug 2018 13:03:14	U	SMPL
J4278-06 - 1	EO-03-080118-C	9	0.0146	-	7470A	PPB	03 Aug 2018 13:05:21	U	SMPL
CCV91 - 1	CCV91	18136	4.9968	-	7470A	PPB	03 Aug 2018 13:07:27	U	SMPL
CCB91 - 1	CCB91	-148	-0.0286	-	7470A	PPB	03 Aug 2018 13:09:34	U	SMPL
J4285-04 - 1	KG-PIT-1	19	0.0173	-	7470A	PPB	03 Aug 2018 13:11:41	U	SMPL
J4285-04DUP - 1	KG-PIT-1DUP	12	0.0154	-	7470A	PPB	03 Aug 2018 13:13:48	U	SMPL
J4285-04MS - 1	KG-PIT-1MS	12657	3.4909	-	7470A	PPB	03 Aug 2018 13:15:55	U	SMPL
J4285-04MSD - 1	KG-PIT-1MSD	13255	3.6553	-	7470A	PPB	03 Aug 2018 13:18:02	U	SMPL
J4285-08 - 1	KG-PIT-2	-103	-0.0162	-	7470A	PPB	03 Aug 2018 13:20:10	U	SMPL
J4285-12 - 1	KG-PIT-3	35	0.0217	-	7470A	PPB	03 Aug 2018 13:22:17	U	SMPL
J4285-16 - 1	KG-PIT-4	64	0.0297	-	7470A	PPB	03 Aug 2018 13:24:25	U	SMPL
J4285-20 - 1	KG-PIT-5	0	0.0121	-	7470A	PPB	03 Aug 2018 13:26:34	U	SMPL
J4285-24 - 1	KG-PIT-6	31	0.0206	-	7470A	PPB	03 Aug 2018 13:28:43	U	SMPL
PB111708BL - 1	PBW	24	0.0187	-	7470A	PPB	03 Aug 2018 13:30:49	U	SMPL
CCV92 - 1	CCV92	18401	5.0697	-	7470A	PPB	03 Aug 2018 13:32:56	U	SMPL
CCB92 - 1	CCB92	-137	-0.0255	-	7470A	PPB	03 Aug 2018 13:35:03	U	SMPL
PB111708BS - 1	LCSW	12735	3.5124	-	7470A	PPB	03 Aug 2018 13:37:11	U	SMPL
PB111708TB - 1	PB111682TB	-143	-0.0272	-	7470A	PPB	03 Aug 2018 13:39:18	U	SMPL
J4284-03 - 1	OILY-WATER	35	0.0217	-	7470A	PPB	03 Aug 2018 13:41:25	U	SMPL
J4284-03DUP - 1	OILY-WATERDUP	-17	0.0074	-	7470A	PPB	03 Aug 2018 13:43:32	U	SMPL
J4284-03MS - 1	OILY-WATERMS	13089	3.6097	-	7470A	PPB	03 Aug 2018 13:45:40	U	SMPL
J4284-03MSD - 1	OILY-WATERMSD	12928	3.5654	-	7470A	PPB	03 Aug 2018 13:47:47	U	SMPL
PB111709BL - 1	PBW	-159	-0.0316	-	7470A	PPB	03 Aug 2018 13:49:55	U	SMPL
PB111709BS - 1	LCSW	13616	3.7545	-	7470A	PPB	03 Aug 2018 13:52:04	U	SMPL
J4262-01 - 1	GW-04	118	0.0446	-	7470A	PPB	03 Aug 2018 13:54:12	U	SMPL
J4262-02 - 1	GW-03	-60	-0.0044	-	7470A	PPB	03 Aug 2018 13:56:19	U	SMPL
CCV93 - 1	CCV93	18207	5.0164	-	7470A	PPB	03 Aug 2018 13:58:26	U	SMPL
CCB93 - 1	CCB93	-159	-0.0316	-	7470A	PPB	03 Aug 2018 14:00:34	U	SMPL
J4262-02DUP - 1	GW-03DUP	-19	0.0069	-	7470A	PPB	03 Aug 2018 14:02:42	U	SMPL
J4262-03 - 1	J4262-02MS	12910	3.5605	-	7470A	PPB	03 Aug 2018 14:04:49	U	SMPL
J4262-04 - 1	J4262-02MSD	13024	3.5918	-	7470A	PPB	03 Aug 2018 14:06:56	U	SMPL
J4262-05 - 1	GW-02	259	0.0833	-	7470A	PPB	03 Aug 2018 14:09:03	U	SMPL
J4262-06 - 1	GW-01	608	0.1792	-	7470A	PPB	03 Aug 2018 14:11:11	U	SMPL
J4262-07 - 1	DUP-073118	-34	0.0028	-	7470A	PPB	03 Aug 2018 14:13:19	U	SMPL
J4262-09 - 1	FB-073118	-74	-0.0082	-	7470A	PPB	03 Aug 2018 14:15:28	U	SMPL
J4279-04 - 1	WE13-GW-TANK1-IDW-072918	-48	-0.0011	-	7470A	PPB	03 Aug 2018 14:17:36	U	SMPL
J4279-05 - 1	WE13-GW-TANK2-IDW-072918	-61	-0.0046	-	7470A	PPB	03 Aug 2018 14:19:45	U	SMPL
J4284-02 - 1	OILY-WATER	747	0.2174	-	7470A	PPB	03 Aug 2018 14:21:52	U	SMPL
CCV94 - 1	CCV94	18188	5.0111	-	7470A	PPB	03 Aug 2018 14:23:59	U	SMPL
CCB94 - 1	CCB94	-157	-0.0310	-	7470A	PPB	03 Aug 2018 14:26:06	U	SMPL
J4287-01 - 1	A4742	371	0.1141	-	7470A	PPB	03 Aug 2018 14:28:14	U	SMPL
J4285-04LX5 - 1		-1	0.0118	-	7470A	PPB	03 Aug 2018 14:30:21	U	SMPL
J4285-04A - 1		13928	3.8403	-	7470A	PPB	03 Aug 2018 14:32:28	U	SMPL
J4284-03LX5 - 1		-117	-0.0200	-	7470A	PPB	03 Aug 2018 14:34:36	U	SMPL
J4284-03A - 1		13890	3.8298	-	7470A	PPB	03 Aug 2018 14:36:43	U	SMPL
J4262-02LX5 - 1		-169	-0.0343	-	7470A	PPB	03 Aug 2018 14:38:51	U	SMPL
J4262-02A - 1		14129	3.8955	-	7470A	PPB	03 Aug 2018 14:40:59	U	SMPL
CCV95 - 1	CCV95	18009	4.9619	-	7470A	PPB	03 Aug 2018 14:43:07	U	SMPL
CCB95 - 1	CCB95	-164	-0.0330	-	7470A	PPB	03 Aug 2018 14:45:17	U	SMPL