



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97101

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	08/03/2018	09:40
2	J4316-10		1	8	>100.0	>212.0	>212.0	08/03/2018	10:15
3	J4316-10DUP		1	8	>100.0	>212.0	>212.0	08/03/2018	10:45
4	J4317-02		1	6	>100.0	>212.0	>212.0	08/03/2018	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB97101

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp8318

WorkList ID : 115245

Date : 8/3/2018 8:58:22 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2018	Water	J4316-10 C	Flash Point	Cool 4 deg C	PSEG01	K63	OW-2-4	08/02/2018	1010A
08/07/2018	Water	J4317-02 C	Flash Point	Cool 4 deg C	PSEG01	K63	N28424	08/02/2018	1010A

Date/Time

Received by:

Relinquished by:

08/03/18 9:05AM

KS

CP

Date/Time

Received by:

Relinquished by:

08/03/18 11:25AM

CP

KS