



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97108

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/03/2018	12:50
2	CAL2	1	Water			21.8	7.01	08/03/2018	12:52
3	CAL3	1	Water			22.0	10.04	08/03/2018	12:54
4	ICV	1	Water			21.9	7.00	08/03/2018	12:56
5	CCV1	1	Water			22.1	2.00	08/03/2018	12:58
6	J4343-01	1	Solid	20.02	20	23.5	6.68	08/03/2018	13:00
7	J4343-01DUP	1	Solid	20.04	20	23.6	6.69	08/03/2018	13:02
8	CCV2	1	Water			21.9	12.00	08/03/2018	13:04

WORKLIST(Hardcopy Internal Chain)

LB97108

WorkList Name : pH080318

WorkList ID : 115263

Date : 8/3/2018 12:34:56 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/02/2018	Solid	J4343-01	pH	Cool 4 deg C	CTIA01	Q22	WJ-18-08-02-TSH9-01	08/02/2018	9045D

Date/Time 08/03/18 12:40
 Received by: SP
 Relinquished by: CP

Date/Time 08/03/18 1:15
 Received by: CP
 Relinquished by: SP