



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97109

Slope : 98.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	08/03/2018	16:00
2	CAL2	1	Water			22.3	7.01	08/03/2018	16:02
3	CAL3	1	Water			22.6	10.04	08/03/2018	16:04
4	ICV	1	Water			22.4	7.00	08/03/2018	16:06
5	CCV1	1	Water			22.3	2.00	08/03/2018	16:08
6	J4352-01	1	Water			23.1	3.37	08/03/2018	16:10
7	J4352-01DUP	1	Water			23.2	3.38	08/03/2018	16:12
8	CCV2	1	Water			22.0	12.00	08/03/2018	16:14

WORKLIST(Hardcopy Internal Chain)

LB97109

WorkList Name : corrosivity080318-1

WorkList ID : 115277

Date : 8/3/2018 3:33:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/03/2018	Water	J4352-01	Corrosivity	Cool 4 deg C	SEAC01	Q32	GROUNDWATER	08/03/2018	9040C

Date/Time 08/03/18 3:45

Received by: SP

Relinquished by: CP

Date/Time 08/03/18 4:45

Received by: CP

Relinquished by: SP