



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97125

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	08/06/2018	11:20
2	CAL2	1	Water			21.9	7.01	08/06/2018	11:22
3	CAL3	1	Water			21.7	10.04	08/06/2018	11:24
4	ICV	1	Water			21.9	7.00	08/06/2018	11:26
5	CCV1	1	Water			21.8	2.00	08/06/2018	11:28
6	J4355-03	1	Water			17.6	10.62	08/06/2018	11:30
7	J4355-03DUP	1	Water			17.7	10.63	08/06/2018	11:32
8	CCV2	1	Water			21.9	12.00	08/06/2018	11:34

LB 97125

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity080618 WorkList ID : 115324 Date : 8/6/2018 11:05:39 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/03/2018	Water	J4355-03	Corrosivity	Cool 4 deg C	BRIN05	Q51	WC-WATER	08/03/2018	9040C

Date/Time 08/06/18 11:10
Received by: Sp
Relinquished by: CP

Date/Time 08/06/18 11:50
Received by: CP
Relinquished by: Sp