



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97131

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.6	08/06/2018	10:10
2	J4352-01		1	7	>100.0	>212.0	>212.0	08/06/2018	10:35
3	J4352-01DUP		1	7	>100.0	>212.0	>212.0	08/06/2018	11:02
4	J4355-03		1	8	>100.0	>212.0	>212.0	08/06/2018	11:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 97131

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP8618

WorkList ID : 115304

Date : 8/6/2018 9:17:42 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/10/2018	Water	J4352-01	Flash Point	Cool 4 deg C	SEAC01	Q32	GROUNDWATER	08/03/2018	1010A
08/17/2018	Water	J4355-03	Flash Point	Cool 4 deg C	BRIN05	Q51	WC-WATER	08/03/2018	1010A

Date/Time 08/06/18 9:30 AM
Received by: KS
Relinquished by: CP

Date/Time 08/06/18 11:40 AM
Received by: CP
Relinquished by: KS