

245.1


LB97132

INSTRUMENT ID: CV2

Linear

 μ Abs.:

34074



A= 0.0000e+000

B= 2.9238e-004

C= -3.0376e-002

Rho= 0.9998042

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	97	0.000	97				
0.05	0.050	0.052	0.002	283	0.0 %	283				
0.2	0.200	0.208	0.008	816	0.0 %	816				
2.5	2.500	2.513	0.013	8698	0.0 %	8698				
5.0	5.000	4.897	-0.103	16852	0.0 %	16852				
7.5	7.500	7.649	0.149	26266	0.0 %	26266				
10.0	10.000	9.932	-0.068	34074	0.0 %	34074				

LB97132 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 06 Aug 2018 14:16:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	97	-	0.0000	245.1	PPB	06 Aug 2018 14:19:58	S	Std
0.05 - 1	50.05	283	-	0.0500	245.1	PPB	06 Aug 2018 14:22:06	S	Std
0.2 - 1	50.2	816	-	0.2000	245.1	PPB	06 Aug 2018 14:24:14	S	Std
2.5 - 1	52.5	8698	-	2.5000	245.1	PPB	06 Aug 2018 14:26:23	S	Std
5.0 - 1	55	16852	-	5.0000	245.1	PPB	06 Aug 2018 14:28:31	S	Std
7.5 - 1	57.5	26266	-	7.5000	245.1	PPB	06 Aug 2018 14:30:37	S	Std
10.0 - 1	510	34074	-	10.0000	245.1	PPB	06 Aug 2018 14:32:46	S	Std
ICV90 - 1	ICV90	13531	3.9259	-245.1		PPB	06 Aug 2018 14:35:22	U	SMPL
ICB90 - 1	ICB90	-125	-0.0669	-245.1		PPB	06 Aug 2018 14:37:30	U	SMPL
CCV09 - 1	CCV09	17465	5.0761	-245.1		PPB	06 Aug 2018 14:39:38	U	SMPL
CCB09 - 1	CCB09	-98	-0.0590	-245.1		PPB	06 Aug 2018 14:41:44	U	SMPL
CRA - 1	CRA	280	0.0515	-245.1		PPB	06 Aug 2018 14:43:53	U	SMPL
HighStd - 1	HighStd	34180	9.9633	-245.1		PPB	06 Aug 2018 14:45:59	U	SMPL
ChkStd - 1	ChkStd	25077	7.3018	-245.1		PPB	06 Aug 2018 14:49:55	U	SMPL
PB111772BL - 1	PBW	19	-0.0248	-245.1		PPB	06 Aug 2018 14:54:28	U	SMPL
PB111772BS - 1	LCSW	13990	4.0601	-245.1		PPB	06 Aug 2018 14:56:36	U	SMPL
J4330-01 - 1	073018-DBRI-E-D-B	-209	-0.0915	-245.1		PPB	06 Aug 2018 14:58:43	U	SMPL
J4330-02 - 1	073018-DBRI-E-D-M	-43	-0.0429	-245.1		PPB	06 Aug 2018 15:00:52	U	SMPL
J4330-03 - 1	073018-DBRI-E-D-S	-41	-0.0424	-245.1		PPB	06 Aug 2018 15:03:00	U	SMPL
J4330-04 - 1	073018-DBRI-E-U-B	-10	-0.0333	-245.1		PPB	06 Aug 2018 15:05:08	U	SMPL
J4330-05 - 1	073018-DBRI-E-U-M	3	-0.0295	-245.1		PPB	06 Aug 2018 15:07:15	U	SMPL
CCV10 - 1	CCV10	17388	5.0536	-245.1		PPB	06 Aug 2018 15:09:21	U	SMPL
CCB10 - 1	CCB10	-109	-0.0622	-245.1		PPB	06 Aug 2018 15:11:28	U	SMPL
J4330-06 - 1	073018-DBRI-E-U-S	-63	-0.0488	-245.1		PPB	06 Aug 2018 15:13:36	U	SMPL
J4330-07 - 1	073018-TIPC-F-D-B	-29	-0.0389	-245.1		PPB	06 Aug 2018 15:15:43	U	SMPL
J4330-08 - 1	073018-TIPC-F-D-S	8	-0.0280	-245.1		PPB	06 Aug 2018 15:17:50	U	SMPL
J4330-09 - 1	073018-TIPC-F-U-B	-15	-0.0348	-245.1		PPB	06 Aug 2018 15:19:57	U	SMPL
J4330-10 - 1	073018-TIPC-F-U-S	-52	-0.0456	-245.1		PPB	06 Aug 2018 15:22:04	U	SMPL
J4331-01 - 1	073118-TIPC-E-D-M	-42	-0.0427	-245.1		PPB	06 Aug 2018 15:24:11	U	SMPL
J4331-02 - 1	073118-TIPC-E-U-M	-40	-0.0421	-245.1		PPB	06 Aug 2018 15:26:20	U	SMPL
J4332-01 - 1	080118-DBRI-E-D-B	-52	-0.0456	-245.1		PPB	06 Aug 2018 15:28:28	U	SMPL
J4332-02 - 1	080118-DBRI-E-D-S	-33	-0.0400	-245.1		PPB	06 Aug 2018 15:30:36	U	SMPL
J4332-03 - 1	080118-DBRI-E-U-B	-13	-0.0342	-245.1		PPB	06 Aug 2018 15:32:43	U	SMPL
CCV11 - 1	CCV11	17118	4.9747	-245.1		PPB	06 Aug 2018 15:34:50	U	SMPL
CCB11 - 1	CCB11	-90	-0.0567	-245.1		PPB	06 Aug 2018 15:36:57	U	SMPL
J4332-04 - 1	080118-DBRI-E-U-S	-48	-0.0444	-245.1		PPB	06 Aug 2018 15:39:06	U	SMPL
J4332-05 - 1	080118-TIPC-F-D-B	6	-0.0286	-245.1		PPB	06 Aug 2018 15:41:13	U	SMPL
J4332-06 - 1	080118-TIPC-F-D-S	-23	-0.0371	-245.1		PPB	06 Aug 2018 15:43:20	U	SMPL
J4332-07 - 1	080118-TIPC-F-U-B	2	-0.0298	-245.1		PPB	06 Aug 2018 15:45:27	U	SMPL
J4332-08 - 1	080118-TIPC-F-U-S	-11	-0.0336	-245.1		PPB	06 Aug 2018 15:47:34	U	SMPL
J4332-08DUP - 1	080118-TIPC-F-U-SDUP	-19	-0.0359	-245.1		PPB	06 Aug 2018 15:49:42	U	SMPL
J4332-08MS - 1	080118-TIPC-F-U-SMS	14917	4.3311	-245.1		PPB	06 Aug 2018 15:51:50	U	SMPL
J4332-08MSD - 1	080118-TIPC-F-U-SMSD	14910	4.3291	-245.1		PPB	06 Aug 2018 15:53:58	U	SMPL
PB111773BL - 1	PBW	-124	-0.0666	-245.1		PPB	06 Aug 2018 15:56:06	U	SMPL
PB111773BS - 1	LCSW	13448	3.9016	-245.1		PPB	06 Aug 2018 15:58:14	U	SMPL
CCV12 - 1	CCV12	17065	4.9592	-245.1		PPB	06 Aug 2018 16:00:21	U	SMPL
CCB12 - 1	CCB12	-130	-0.0684	-245.1		PPB	06 Aug 2018 16:02:29	U	SMPL
J4292-01 - 1	COMP	200	0.0281	-245.1		PPB	06 Aug 2018 16:04:38	U	SMPL
J4292-01DUP - 1	COMP DUP	90	-0.0041	-245.1		PPB	06 Aug 2018 16:06:45	U	SMPL
J4292-01MS - 1	COMPMS	12690	3.6800	-245.1		PPB	06 Aug 2018 16:08:52	U	SMPL
J4292-01MSD - 1	COMPMSD	13299	3.8581	-245.1		PPB	06 Aug 2018 16:10:59	U	SMPL
J4352-01 - 1	GROUNDWATER	-108	-0.0620	-245.1		PPB	06 Aug 2018 16:13:06	U	SMPL
J4332-08LX5 - 1		74	-0.0087	-245.1		PPB	06 Aug 2018 16:15:15	U	SMPL
J4332-08A - 1		14001	4.0633	-245.1		PPB	06 Aug 2018 16:17:23	U	SMPL
J4292-01LX5 - 1		-16	-0.0351	-245.1		PPB	06 Aug 2018 16:19:31	U	SMPL
J4292-01A - 1		13810	4.0075	-245.1		PPB	06 Aug 2018 16:21:40	U	SMPL
CCV13 - 1	CCV13	17358	5.0448	-245.1		PPB	06 Aug 2018 16:23:47	U	SMPL
CCB13 - 1	CCB13	-79	-0.0535	-245.1		PPB	06 Aug 2018 16:25:55	U	SMPL