

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/6/2018 15:19

Test: Total CN *My/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	100.171	0.0	0.093	
ICB1	0.285	0.0	0.004	
CCV1	247.681	0.0	0.224	
CCB1	0.099	0.0	0.004	
PB111805BL	-0.010	0.0	0.004	
PB111805BS	101.209	0.0	0.094	
MIDPB111805	249.185	0.0	0.225	
J4329-01	33.792	0.0	0.034	
J4329-01DUP	33.701	0.0	0.034	
J4329-01MS	68.760	0.0	0.065	
J4329-01MSD	68.820	0.0	0.065	
CCV2	261.545	0.0	0.236	
CCB2	-0.092	0.0	0.004	

N	13
Mean	89.627
SD	99.7647
CV%	111.31

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/6/2018 15:24


Test Total CN

Accepted 8/6/2018 8:05

Factor 1126
Bias 0.004

Coeff. of det. 0.999986

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.7431	0.0000	
2	5.0PPBCN	0.008	4.6402	5.0000	
3	10PPBCN	0.012	9.6796	10.0000	
4	50PPBCN	0.049	50.6098	50.0000	
5	100PPBCN	0.094	101.2701	100.0000	
6	250PPBCN	0.226	249.6965	250.0000	
7	500PPBCN	0.448	499.8467	500.0000	

Aquakem v. 7.2AQ1

AS

Results from time period:

Mon Aug 06 15:15:08 2018

Mon Aug 06 15:18:32 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.7431	µg/l	8/6/2018 8:03:28	
5.0PPBCN	A	Total CN	P	4.6402	µg/l	8/6/2018 8:03:29	
10PPBCN	A	Total CN	P	9.6796	µg/l	8/6/2018 8:03:30	
50PPBCN	A	Total CN	P	50.6098	µg/l	8/6/2018 8:03:31	
100PPBCN	A	Total CN	P	101.2701	µg/l	8/6/2018 8:03:32	
250PPBCN	A	Total CN	P	249.6965	µg/l	8/6/2018 8:03:33	
500PPBCN	A	Total CN	P	499.8467	µg/l	8/6/2018 8:03:34	
ICV1	S	Total CN	P	100.1712	µg/l	8/6/2018 15:15:08	
ICB1	S	Total CN	P	0.2849	µg/l	8/6/2018 15:15:09	
CCV1	S	Total CN	P	247.6811	µg/l	8/6/2018 15:15:10	
CCB1	S	Total CN	P	0.0991	µg/l	8/6/2018 15:15:11	
PB111805BL	S	Total CN	P	-0.01	µg/l	8/6/2018 15:15:12	
PB111805BS	S	Total CN	P	101.2089	µg/l	8/6/2018 15:15:13	
MIDPB111805	S	Total CN	P	249.1851	µg/l	8/6/2018 15:15:14	
J4329-01	S	Total CN	P	33.792	µg/l	8/6/2018 15:15:15	
J4329-01DUP	S	Total CN	P	33.7012	µg/l	8/6/2018 15:15:16	
J4329-01MS	S	Total CN	P	68.7604	µg/l	8/6/2018 15:15:17	
J4329-01MSD	S	Total CN	P	68.8195	µg/l	8/6/2018 15:15:18	
CCV2	S	Total CN	P	261.5446	µg/l	8/6/2018 15:18:31	
CCB2	S	Total CN	P	-0.0916	µg/l	8/6/2018 15:18:32	