


7471B

LB97143
INSTRUMENT ID: CV1

Linear

 μ Abs.:

4641E



A= 0.0000e+000

B= 2.1585e-004

C= 1.1337e-002

Rho= 0.9999129

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	139	0.000	139				
0.2	0.200	0.230	0.030	1012	0.0 %	1012				
2.5	2.500	2.402	-0.098	11074	0.0 %	11074				
5.0	5.000	5.022	0.022	23212	0.0 %	23212				
7.5	7.500	7.476	-0.024	34582	0.0 %	34582				
10.0	10.000	10.030	0.030	46416	0.0 %	46416				

LB97143 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Aug 2018 16:45:01

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	139	-	0.00007471B	PPB	06 Aug 2018 16:50:20	S	Std
0.2 - 1	50.2	1012	-	0.20007471B	PPB	06 Aug 2018 16:52:37	S	Std
2.5 - 1	52.5	11074	-	2.50007471B	PPB	06 Aug 2018 16:54:55	S	Std
5.0 - 1	55	23212	-	5.00007471B	PPB	06 Aug 2018 16:57:12	S	Std
7.5 - 1	57.5	34582	-	7.50007471B	PPB	06 Aug 2018 16:59:29	S	Std
10.0 - 1	510	46416	-	10.00007471B	PPB	06 Aug 2018 17:04:21	S	Std
ICV91 - 1	ICV91	17615	3.8135	-7471B	PPB	06 Aug 2018 17:07:02	U	SMPL
ICB91 - 1	ICB91	81	0.0288	-7471B	PPB	06 Aug 2018 17:09:18	U	SMPL
CCV14 - 1	CCV14	23620	5.1096	-7471B	PPB	06 Aug 2018 17:11:35	U	SMPL
CCB14 - 1	CCB14	97	0.0323	-7471B	PPB	06 Aug 2018 17:13:51	U	SMPL
CRA - 1	CRA	1026	0.2328	-7471B	PPB	06 Aug 2018 17:16:09	U	SMPL
HighStd - 1	HighStd	47111	10.1800	-7471B	PPB	06 Aug 2018 17:18:25	U	SMPL
ChkStd - 1	ChkStd	32994	7.1329	-7471B	PPB	06 Aug 2018 17:20:40	U	SMPL
PB111774BL - 1	PBS	-145	-0.0200	-7471B	PPB	06 Aug 2018 17:22:59	U	SMPL
PB111774BS - 1	LCSS	15402	3.3358	-7471B	PPB	06 Aug 2018 17:25:16	U	SMPL
J4277-01 - 1	OR-02-080318-A	2399	0.5292	-7471B	PPB	06 Aug 2018 17:27:33	U	SMPL
J4277-03 - 1	OR-02-080318-B	4565	0.9967	-7471B	PPB	06 Aug 2018 17:29:50	U	SMPL
J4283-01 - 1	TR-04-080318	1386	0.3105	-7471B	PPB	06 Aug 2018 17:32:08	U	SMPL
J4327-01 - 1	FK-BPM-02-007-ABCD	1948	0.4318	-7471B	PPB	06 Aug 2018 17:34:27	U	SMPL
J4327-01DUP - 1	FK-BPM-02-007-ABCD DUP	1980	0.4387	-7471B	PPB	06 Aug 2018 17:36:42	U	SMPL
CCV15 - 1	CCV15	23195	5.0179	-7471B	PPB	06 Aug 2018 17:38:58	U	SMPL
CCB15 - 1	CCB15	58	0.0239	-7471B	PPB	06 Aug 2018 17:41:14	U	SMPL
J4327-01MS - 1	FK-BPM-02-007-ABCDMS	18624	4.0312	-7471B	PPB	06 Aug 2018 17:43:30	U	SMPL
J4327-01MSD - 1	FK-BPM-02-007-ABCDMSD	18576	4.0209	-7471B	PPB	06 Aug 2018 17:45:46	U	SMPL
J4335-01 - 1	CINNAMINSON-8-3-18	5507	1.2000	-7471B	PPB	06 Aug 2018 17:48:03	U	SMPL
J4343-01 - 1	WJ-18-08-02-TSH9-01	2502	0.5514	-7471B	PPB	06 Aug 2018 17:50:20	U	SMPL
J4345-01 - 1	RBR200035	1660	0.3696	-7471B	PPB	06 Aug 2018 17:52:36	U	SMPL
J4347-01 - 1	CONCRETE-PILE-1	1662	0.3701	-7471B	PPB	06 Aug 2018 17:54:53	U	SMPL
J4348-01 - 1	72-11930	6946	1.5106	-7471B	PPB	06 Aug 2018 17:57:10	U	SMPL
J4355-01 - 1	WC-SOIL	419	0.1018	-7471B	PPB	06 Aug 2018 17:59:28	U	SMPL
J4327-01LX5 - 1		446	0.1076	-7471B	PPB	06 Aug 2018 18:01:45	U	SMPL
J4327-01A - 1		20611	4.4601	-7471B	PPB	06 Aug 2018 18:04:02	U	SMPL
CCV16 - 1	CCV16	23190	5.0168	-7471B	PPB	06 Aug 2018 18:06:18	U	SMPL
CCB16 - 1	CCB16	41	0.0202	-7471B	PPB	06 Aug 2018 18:08:34	U	SMPL