



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97148

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	08/07/2018	09:40
2	J4366-02		1	8	>100.0	>212.0	>212.0	08/07/2018	10:10
3	J4366-02DUP		1	8	>100.0	>212.0	>212.0	08/07/2018	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP8718

WorkList ID : 115357

Date : 8/7/2018 9:08:54 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/09/2018	Water	J4366-02	C Flash Point	Cool 4 deg C	PSEG01	K63	SV30518L	08/06/2018	1010A

LB 97148

Date/Time 08/07/18 09:15 AM
 Received by: KS
 Relinquished by: CP

Date/Time 08/07/18 10:50 AM
 Received by: CP
 Relinquished by: KS