

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/7/2018 11:10

Test: Phenolics-*mg/L*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.065	0.0	0.111	
ICB1	0.001	0.0	0.004	
CCV1	1.067	0.0	0.111	
CCB1	-0.002	0.0	0.004	
PB111811BL	-0.000	0.0	0.004	
PB111811BS	1.071	0.0	0.111	
J4352-01	0.003	0.0	0.004	
J4352-01DUP	0.002	0.0	0.004	
J4352-01MS	0.878	0.0	0.092	
J4352-01MSD	0.861	0.0	0.090	
CCV2	0.996	0.0	0.104	
CCB2	-0.001	0.0	0.004	

N	12
Mean	0.495
SD	0.5207
CV%	105.15

Aquakem v. 7.2AQ1

Results from time period:

Tue Aug 07 09:53:57 2018

Tue Aug 07 11:07:59 2018

AS

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.052	mg/l	8/7/2018 9:53:57	
0.1PPM	A	Phenolics-	P	0.0993	mg/l	8/7/2018 9:53:58	
0.25PPM	A	Phenolics-	P	0.2535	mg/l	8/7/2018 9:53:59	
0.50PPM	A	Phenolics-	P	0.4955	mg/l	8/7/2018 9:54:00	
1.0PPM	A	Phenolics-	P	0.998	mg/l	8/7/2018 9:54:01	
2.0PPM	A	Phenolics-	P	2.0017	mg/l	8/7/2018 9:54:02	
ICV1	S	Phenolics-	P	1.0648	mg/l	8/7/2018 10:43:46	
ICB1	S	Phenolics-	P	0.0013	mg/l	8/7/2018 10:43:47	
CCV1	S	Phenolics-	P	1.0674	mg/l	8/7/2018 11:07:48	
CCB1	S	Phenolics-	P	-0.002	mg/l	8/7/2018 11:07:49	
PB111811BL	S	Phenolics-	P	-0.0003	mg/l	8/7/2018 11:07:50	
PB111811BS	S	Phenolics-	P	1.0708	mg/l	8/7/2018 11:07:51	
J4352-01	S	Phenolics-	P	0.0032	mg/l	8/7/2018 11:07:52	
J4352-01DUP	S	Phenolics-	P	0.0024	mg/l	8/7/2018 11:07:53	
J4352-01MS	S	Phenolics-	P	0.8784	mg/l	8/7/2018 11:07:54	
J4352-01MSD	S	Phenolics-	P	0.8611	mg/l	8/7/2018 11:07:55	
CCV2	S	Phenolics-	P	0.996	mg/l	8/7/2018 11:07:56	
CCB2	S	Phenolics-	P	-0.0008	mg/l	8/7/2018 11:07:57	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/7/2018 10:21


Test Phenolics-

Accepted 8/7/2018 10:21

Factor 9.961
Bias 0.004

Coeff. of det. 0.999984

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0520	0.0500	
2	Phenol-2	0.014	0.0993	0.1000	
3	Phenol-2	0.029	0.2535	0.2500	
4	Phenol-2	0.054	0.4955	0.5000	
5	Phenol-2	0.104	0.9980	1.0000	
6	Phenol-2	0.205	2.0017	2.0000	