

Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB97152

Reagent/Standard	Lot/Loq #
MBAS CAL STD, 0.5PPM	WP66789
MBAS CCV STD, 1.0PPM	WP66799
MBAS ICV-LCS std 1.0PPM	WP66796
MBAS DAILY CHECK STD, 0.01PPM	WP66797
MBAS DAILY CHECK STD, 2.0PPM	WP66798
MBAS CAL STD 2 PPM	WP66792
MBAS CAL STD 1 PPM	WP66790
MBAS CAL STD 0.1 PPM	WP66786
mbas cal std 0.05 ppm	WP66794
MBAS CAL STD 0.01 PPM	WP66793
MBAS BLANK 0 PPM	WP66784
MBAS CAL STD 0.4 PPM	WP66788
MBAS CAL STD 0.2 PPM	WP66787
Methylene blue reagent-1000ml	WP66778
4-AMINOANTIPYRINE	WP66779

Intercept: 0.0008

Slope: 0.2214

Regression: 0.9999

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Anal Date	Anal Time
1	CAL1	0	1	0.000	08/07/2018	11:07
2	CAL2	0.01	1	0.003	08/07/2018	11:08
3	CAL3	0.05	1	0.011	08/07/2018	11:08
4	CAL4	0.1	1	0.023	08/07/2018	11:09
5	CAL5	0.2	1	0.043	08/07/2018	11:09
6	CAL6	0.4	1	0.088	08/07/2018	11:10
7	CAL7	0.5	1	0.116	08/07/2018	11:10
8	CAL8	1	1	0.224	08/07/2018	11:11
9	CAL9	2	1	0.442	08/07/2018	11:11



Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB97152

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Result (mg/L)	AnalDate	Anal Time
1	ICV	1	1	0.210	0.945	08/07/2018	11:12
2	ICB		1	0.000	-0.004	08/07/2018	11:12
3	CHK1	0.01	1	0.003	0.010	08/07/2018	11:13
4	CHK2	2	1	0.408	1.839	08/07/2018	11:14
5	CCV1	1	1	0.218	0.981	08/07/2018	11:14
6	CCB1		1	0.000	-0.004	08/07/2018	11:15
7	LB97152BL		1	0.000	-0.004	08/07/2018	11:15
8	LB97152BS	1	1	0.210	0.945	08/07/2018	11:15
9	J4080-26		1	0.171	0.769	08/07/2018	11:16
10	J4080-26DUP		1	0.172	0.773	08/07/2018	11:16
11	CCV2	1	1	0.221	0.995	08/07/2018	11:17
12	CCB2		1	0.000	-0.004	08/07/2018	11:18