

245.1


LB 97157

INSTRUMENT ID: CV2

Linear

 μ Abs.:

31774



A= 0.0000e+000

B= 3.0777e-004

C= -1.4443e-002

Rho= 0.9992823

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	-17	0.000	-17				
0.05	0.050	0.049	-0.001	207	0.0 %	207				
0.2	0.200	0.181	-0.019	635	0.0 %	635				
2.5	2.500	2.421	-0.079	7914	0.0 %	7914				
5.0	5.000	5.094	0.094	16599	0.0 %	16599				
7.5	7.500	7.759	0.259	25258	0.0 %	25258				
10.0	10.000	9.765	-0.235	31774	0.0 %	31774				

LB97157 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Aug 2018 15:25:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-17	-	0.0000	245.1	PPB	07 Aug 2018 15:27:52	S	Std
0.05 - 1	50.05	207	-	0.0500	245.1	PPB	07 Aug 2018 15:30:00	S	Std
0.2 - 1	50.2	635	-	0.2000	245.1	PPB	07 Aug 2018 15:32:08	S	Std
2.5 - 1	52.5	7914	-	2.5000	245.1	PPB	07 Aug 2018 15:34:16	S	Std
5.0 - 1	55	16599	-	5.0000	245.1	PPB	07 Aug 2018 15:36:24	S	Std
7.5 - 1	57.5	25258	-	7.5000	245.1	PPB	07 Aug 2018 15:38:31	S	Std
10.0 - 1	510	31774	-	10.0000	245.1	PPB	07 Aug 2018 15:47:16	S	Std
ICV93 - 1	ICV93	13176	4.0407	-245.1		PPB	07 Aug 2018 15:50:59	U	SMPL
ICB93 - 1	ICB93	-182	-0.0705	-245.1		PPB	07 Aug 2018 15:53:06	U	SMPL
CCV22 - 1	CCV22	17087	5.2444	-245.1		PPB	07 Aug 2018 15:55:13	U	SMPL
CCB22 - 1	CCB22	-176	-0.0686	-245.1		PPB	07 Aug 2018 15:57:19	U	SMPL
CRA - 1	CRA	175	0.0394	-245.1		PPB	07 Aug 2018 15:59:27	U	SMPL
HighStd - 1	HighStd	31533	9.6905	-245.1		PPB	07 Aug 2018 16:01:33	U	SMPL
ChkStd - 1	ChkStd	22921	7.0400	-245.1		PPB	07 Aug 2018 16:03:40	U	SMPL
PB111808BL - 1	PBW	322	0.0847	-245.1		PPB	07 Aug 2018 16:08:31	U	SMPL
PB111808BS - 1	LCSW	14450	4.4328	-245.1		PPB	07 Aug 2018 16:10:38	U	SMPL
J4350-01 - 1	080218-DBRI-E-D-B	609	0.1730	-245.1		PPB	07 Aug 2018 16:12:45	U	SMPL
J4350-02 - 1	080218-DBRI-E-D-S	146	0.0305	-245.1		PPB	07 Aug 2018 16:14:53	U	SMPL
J4350-03 - 1	080218-DBRI-E-U-B	119	0.0222	-245.1		PPB	07 Aug 2018 16:17:01	U	SMPL
J4350-04 - 1	080218-DBRI-E-U-S	120	0.0225	-245.1		PPB	07 Aug 2018 16:19:10	U	SMPL
J4350-05 - 1	080218-TIPC-E-D-M	145	0.0302	-245.1		PPB	07 Aug 2018 16:21:16	U	SMPL
CCV23 - 1	CCV23	17131	5.2580	-245.1		PPB	07 Aug 2018 16:23:23	U	SMPL
CCB23 - 1	CCB23	-181	-0.0701	-245.1		PPB	07 Aug 2018 16:25:30	U	SMPL
J4350-06 - 1	080218-TIPC-E-U-M	388	0.1050	-245.1		PPB	07 Aug 2018 16:27:38	U	SMPL
J4350-06DUP - 1	080218-TIPC-E-U-MDUP	348	0.0927	-245.1		PPB	07 Aug 2018 16:38:56	U	SMPL
J4350-06MS - 1	080218-TIPC-E-U-MMS	16120	4.9468	-245.1		PPB	07 Aug 2018 16:41:03	U	SMPL
J4350-06MSD - 1	080218-TIPC-E-U-MMSD	16119	4.9465	-245.1		PPB	07 Aug 2018 16:43:10	U	SMPL
J4350-06LX5 - 1		59	0.0037	-245.1		PPB	07 Aug 2018 16:47:26	U	SMPL
J4350-06A - 1		13395	4.1082	-245.1		PPB	07 Aug 2018 16:49:35	U	SMPL
CCV24 - 1	CCV24	16680	5.1192	-245.1		PPB	07 Aug 2018 16:51:43	U	SMPL
CCB24 - 1	CCB24	-189	-0.0726	-245.1		PPB	07 Aug 2018 16:53:51	U	SMPL