



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97158

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/07/2018	13:20
2	CAL2	1	Water			22.0	7.01	08/07/2018	13:22
3	CAL3	1	Water			22.4	10.04	08/07/2018	13:24
4	ICV	1	Water			21.9	7.00	08/07/2018	13:26
5	CCV1	1	Water			22.2	2.00	08/07/2018	13:28
6	J4374-01	1	Water			28.0	6.68	08/07/2018	13:30
7	J4374-01DUP	1	Water			28.1	6.69	08/07/2018	13:32
8	J4374-04	1	Water			31.2	6.67	08/07/2018	13:36
9	J4374-05	1	Water			30.6	6.75	08/07/2018	13:38
10	J4374-06	1	Water			30.1	6.83	08/07/2018	13:40
11	J4374-07	1	Water			30.4	6.84	08/07/2018	13:42
12	J4374-08	1	Water			29.7	6.73	08/07/2018	13:44
13	J4374-09	1	Water			29.8	6.84	08/07/2018	13:46
14	J4374-10	1	Water			29.9	6.76	08/07/2018	13:48
15	J4374-11	1	Water			29.4	6.78	08/07/2018	13:50
16	CCV2	1	Water			21.9	12.00	08/07/2018	13:52
17	J4374-12	1	Water			29.6	6.87	08/07/2018	13:54
18	CCV3	1	Water			21.8	2.00	08/07/2018	13:56

WORKLIST(Hardcopy Internal Chain)

LB 97158

WorkList Name : pH080718

WorkList ID : 115364

Date : 8/7/2018 12:56:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2018	Water	J4374-01	pH	Cool 4 deg C	PSEG01	K63	OILY-WATER	08/07/2018	9040C
08/07/2018	Water	J4374-04	pH	Cool 4 deg C	PSEG01	K63	DRUM-1	08/07/2018	9040C
08/07/2018	Water	J4374-05	pH	Cool 4 deg C	PSEG01	K63	DRUM-2	08/07/2018	9040C
08/07/2018	Water	J4374-06	pH	Cool 4 deg C	PSEG01	K63	DRUM-3	08/07/2018	9040C
08/07/2018	Water	J4374-07	pH	Cool 4 deg C	PSEG01	K63	DRUM-4	08/07/2018	9040C
08/07/2018	Water	J4374-08	pH	Cool 4 deg C	PSEG01	K63	DRUM-5	08/07/2018	9040C
08/07/2018	Water	J4374-09	pH	Cool 4 deg C	PSEG01	K63	DRUM-6	08/07/2018	9040C
08/07/2018	Water	J4374-10	pH	Cool 4 deg C	PSEG01	K63	DRUM-7	08/07/2018	9040C
08/07/2018	Water	J4374-11	pH	Cool 4 deg C	PSEG01	K63	DRUM-8	08/07/2018	9040C
08/07/2018	Water	J4374-12	pH	Cool 4 deg C	PSEG01	K63	DRUM-9	08/07/2018	9040C

Date/Time 08/07/18 12:10
 Received by: SP
 Relinquished by: CP

Date/Time 08/07/18 2:10
 Received by: CP
 Relinquished by: SP