



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB97159

Analyst By : sweta

Supervisor Review By : apatel

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/07/2018	13:20
2	CAL2	1	Water			22.0	7.01	08/07/2018	13:22
3	CAL3	1	Water			22.4	10.04	08/07/2018	13:24
4	ICV	1	Water			21.9	7.00	08/07/2018	13:26
5	CCV1	1	Water			22.2	2.00	08/07/2018	13:28
6	J4374-02	1	Water			28.0	6.68	08/07/2018	13:30
7	J4374-02DUP	1	Water			28.1	6.69	08/07/2018	13:32
8	CCV2	1	Water			21.9	12.00	08/07/2018	13:52

LB97159

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity080718

WorkList ID : 115365

Date : 8/7/2018 12:56:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2018	Water	J4374-02	B Corrosivity	Cool 4 deg C	PSEG01	K63	OILY-WATER	08/07/2018	9040C

Date/Time 08/07/18 1:10
Received by: SP
Relinquished by: SP

08/07/18
Date/Time 08/08/18 2:10
Received by: SP
Relinquished by: SP