


7471B

LB97164
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43333



A= 0.0000e+000

B= 2.3070e-004

C= -7.0710e-002

Rho= 0.9999113

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	339	0.000	339				
0.2	0.200	0.160	-0.040	998	0.0 %	998				
2.5	2.500	2.488	-0.012	11092	0.0 %	11092				
5.0	5.000	5.051	0.051	22202	0.0 %	22202				
7.5	7.500	7.568	0.068	33110	0.0 %	33110				
10.0	10.000	9.926	-0.074	43333	0.0 %	43333				

LB97164 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Aug 2018 17:22:39

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SO	339	-	0.00007471B	PPB	07 Aug 2018 17:23:22	S	Std
0.2 - 1	SO.2	998	-	0.20007471B	PPB	07 Aug 2018 17:25:41	S	Std
2.5 - 1	SO.5	11092	-	2.50007471B	PPB	07 Aug 2018 17:27:59	S	Std
5.0 - 1	SO.5	22202	-	5.00007471B	PPB	07 Aug 2018 17:30:16	S	Std
7.5 - 1	SO.5	33110	-	7.50007471B	PPB	07 Aug 2018 17:32:34	S	Std
10.0 - 1	SO.5	43333	-	10.00007471B	PPB	07 Aug 2018 17:34:52	S	Std
ICV94 - 1	ICV94	16657	3.7720	-7471B	PPB	07 Aug 2018 17:39:49	U	SMPL
ICB94 - 1	ICB94	71	-0.0543	-7471B	PPB	07 Aug 2018 17:42:04	U	SMPL
CCV25 - 1	CCV25	22514	5.1232	-7471B	PPB	07 Aug 2018 17:44:20	U	SMPL
CCB25 - 1	CCB25	28	-0.0643	-7471B	PPB	07 Aug 2018 17:46:36	U	SMPL
CRA - 1	CRA	923	0.1422	-7471B	PPB	07 Aug 2018 17:48:53	U	SMPL
HighStd - 1	HighStd	44418	10.1763	-7471B	PPB	07 Aug 2018 17:51:08	U	SMPL
ChkStd - 1	ChkStd	31280	7.1454	-7471B	PPB	07 Aug 2018 17:53:24	U	SMPL
PB111821BL - 1	PBS	-81	-0.0894	-7471B	PPB	07 Aug 2018 17:55:42	U	SMPL
PB111821BS - 1	LCSS	16194	3.6652	-7471B	PPB	07 Aug 2018 17:58:01	U	SMPL
J4360-01 - 1	155-SOIL	3592	0.7579	-7471B	PPB	07 Aug 2018 18:05:14	U	SMPL
J4360-01DUP - 1	155-SOILDUP	3388	0.7109	-7471B	PPB	07 Aug 2018 18:07:31	U	SMPL
J4360-01MS - 1	155-SOILMS	20059	4.5568	-7471B	PPB	07 Aug 2018 18:09:48	U	SMPL
J4360-01MSD - 1	155-SOILMSD	19907	4.5217	-7471B	PPB	07 Aug 2018 18:12:06	U	SMPL
J4360-04 - 1	155-CONCRETE	1167	0.1985	-7471B	PPB	07 Aug 2018 18:14:23	U	SMPL
CCV26 - 1	CCV26	21866	4.9737	-7471B	PPB	07 Aug 2018 18:16:40	U	SMPL
CCB26 - 1	CCB26	94	-0.0490	-7471B	PPB	07 Aug 2018 18:18:56	U	SMPL
J4360-01LX5 - 1		1052	0.1720	-7471B	PPB	07 Aug 2018 18:21:12	U	SMPL
J4360-01A - 1		22015	5.0080	-7471B	PPB	07 Aug 2018 18:23:28	U	SMPL
CCV27 - 1	CCV27	21781	4.9541	-7471B	PPB	07 Aug 2018 18:25:44	U	SMPL
CCB27 - 1	CCB27	141	-0.0382	-7471B	PPB	07 Aug 2018 18:28:00	U	SMPL